


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 221118A/N-111	Page	: 1 of 1	
CALIBRATED ON.	: 19 / 11 / 2022	Ref. Std.	: IS:12508:1988	
CALIBRATION DUE.	: 18 / 11 / 2023	C.P. No.	: ACS/CP/10	
SRF No. & DATE	: 221118A/N Dt. 18.11.2022	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342722000000975F Dt. 22.11.2022	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Plot No. 402, Mogappair, Near Dav Girls School, Paneer Nagar, Chennai			
ITEM CALIBRATED	: Centrifuge			
	Make	: Feb/22	Range	: (0 to 6000) RPM
	L. Count	:	ID No.	: RCL/BIO/Cent-01
	Location	: Clinical Biochemistry	S. No.	: ZIBN-08845
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C} R.H. {40% to 60%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1	11/09/2023
RESULTS	: Mechanical Discipline			
	* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
	1000	1000.96	- 0.96	
	2000	2001.76	- 1.76	
	3000	3002.68	- 2.68	
	4000	4003.44	- 3.44	
	(average of readings)			

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 2.51 rpm of reading at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

