


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 221118A/N-112	Page : 1 of 1			
CALIBRATED ON. : 19 / 11 / 2022	Ref. Std. : IS:12508:1988			
CALIBRATION DUE. : 18 / 11 / 2023	C.P. No. : ACS/CP/10			
SRF No. & DATE : 221118A/N Dt. 18.11.2022	Phy. Inspection : O.K.			
ULR No. & ISSUE DATE : CC342722000000976F Dt. 22.11.2022	Cal. At : Site / Lab			
CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd. Plot No. 402, Mogappair, Near Dav Girls School, Paneer Nagar, Chennai				
ITEM CALIBRATED : Centrifuge				
Make : Feb/22	Range : (0 to 6000) RPM			
L. Count :	ID No. : RCL/CP/Cent-02			
Location : Clinical Pathology	S. No. : ZIBN-08841			
ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. {40% to 60%}				
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1	11/09/2023
RESULTS : Mechanical Discipline				
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)		
1000	1001.32	- 1.32		
2000	2002.12	- 2.12		
3000	3003.20	- 3.20		
4000	4004.24	- 4.24		
(average of readings)				

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 2.52 rpm of reading at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

