



VIEW LAB MONTHLY SUMMARY

Lab Name	GURU MICRO LABORATORY	Details About Robust Analysis
Lab No	16065	Detail About Monthly Summary
Month	November	Detail about SDI
Year	2022	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 18/11/2022

Date of Report Published : 06/12/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Agape	348	187.48	8.36	15.67	191.3 mg/dL	0.24	1.68
2	UREA II	Urease UV / GLDH II	Agape	308	51.71	10.04	5.19	46.9 mg/dL	-0.93	0.59
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Agape	249	2.36	8.64	0.20	2 mg/dL	-1.76	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	256	3.02	11.11	0.34	3.4 mg/dL	1.13	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	260	5.55	7.98	0.44	4.9 g/dL	-1.47	0.05
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	255	3.32	9.76	0.32	2.9 g/dL	-1.30	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	263	5.22	15.89	0.83	5 mg/dL	-0.27	0.10
8	CHOLESTROL II	CHOD-PAP II	Agape	341	107.20	12.47	13.37	113 mg/dL	0.43	1.45
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	240	174.08	9.80	17.06	172 mg/dL	-0.12	2.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out