



VIEW LAB MONTHLY SUMMARY

Lab Name GURU MICRO LABORATORY
 Lab No 16065
 Month October
 Year 2022
 Constituent Group Chemistry II

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/10/2022

Date of Report Published : 03/11/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Agape	354	353.73	9.45	33.42	334.2 mg/dL	-0.58	3.55
2	UREA II	Urease UV / GLDH II	Agape	310	34.21	11.81	4.04	38 mg/dL	0.94	0.46
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Agape	257	1.30	11.53	0.15	1.3 mg/dL	0.00	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	262	1.45	16.49	0.24	1.4 mg/dL	-0.21	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	231	5.55	7.84	0.44	5.9 g/dL	0.80	0.06
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	244	3.27	9.14	0.30	2.9 g/dL	-1.24	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	252	7.06	9.88	0.70	6.8 mg/dL	-0.37	0.09
8	CHOLESTROL II	CHOD-PAP II	Agape	331	112.74	10.03	11.31	106 mg/dL	-0.60	1.24
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	239	228.17	10.10	23.04	248 mg/dL	0.86	2.98

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out