



CHRISTIAN MEDICAL COLLEGE

VELLORE

In Service of the Nation Since 1900...



VIEW LAB MONTHLY SUMMARY

Lab Name **GURU MICRO LABORATORY**

[Details About Robust Analysis](#)

Lab No **16065**

[Detail About Monthly Summary](#)

Month **August**

[Detail about SDI](#)

Year **2022**

Constituent Group **Chemistry II**

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/08/2022

Date of Report Published : 05/09/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Agape	349	112.63	8.54	9.62	122.9 mg/dL	1.07	1.03
2	UREA II	Urease UV / GLDH II	Agape	292	23.73	11.93	2.83	26.2 mg/dL	0.87	0.33
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Agape	236	5.68	10.45	0.59	6.9 mg/dL	2.06	0.08
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	270	4.84	11.62	0.56	5.1 mg/dL	0.46	0.07
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	251	5.74	8.33	0.48	5.6 g/dL	-0.29	0.06
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	242	3.48	9.61	0.33	3.48 g/dL	0.00	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	194	3.57	16.46	0.59	3.39 mg/dL	-0.31	0.08
8	CHOLESTROL II	CHOD-PAP II	Agape	320	111.79	10.25	11.46	120.8 mg/dL	0.79	1.28
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	233	185.99	9.86	18.34	184 mg/dL	-0.11	2.40

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out