



CHRISTIAN MEDICAL COLLEGE

VELLORE

In Service of the Nation Since 1900...



VIEW LAB MONTHLY SUMMARY

Lab Name GURU MICRO LABORATORY
Lab No 16065
Month July
Year 2022
Constituent Group Chemistry II

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/07/2022

Date of Report Published : 05/08/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Agape	331	308.94	8.38	25.90	322.9 mg/dL	0.54	2.85
2	UREA II	Urease UV / GLDH II	Agape	307	58.60	10.43	6.11	47.3 mg/dL	-1.85	0.70
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Agape	247	3.00	9.56	0.29	2.81 mg/dL	-0.66	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	249	1.82	16.24	0.30	2.5 mg/dL	2.30	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	256	5.72	8.09	0.46	5.12 g/dL	-1.30	0.06
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	256	3.44	8.65	0.30	2.91 g/dL	-1.78	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	270	7.48	10.66	0.80	7.46 mg/dL	-0.03	0.10
8	CHOLESTROL II	CHOD-PAP II	Agape	343	114.83	11.62	13.35	95 mg/dL	-1.49	1.44
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	225	266.26	10.57	28.14	258 mg/dL	-0.29	3.75

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out