



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/22/11863		MEDICAL DEVICES	Page No: 1 of 1
Issue Date	29-10-2022		
SRF No & Date	SRF/22/00259-0002 & 28-10-2022		
Receipt Date	28-10-2022		
Calibration Date	28-10-2022		
Calibration Due	27-10-2023		

Customer Name & Address
GOVERNMENT COMMUNITY HEALTH CENTRE,
SARCARSAMAKULAM, COIMBATORE-641107.

Details of Device Under Calibration (DUC)			
Description	ELECTRICAL SAFETY (MICROSCOPE)	Make & Model	LABOMED & VISION 2000
Range		Sr. No	2020
Resolution		Identification No	EQ/LAB/SSK-005
DUC Condition	SATISFACTORY	Location	LAB


Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature: 25.3° C	Relative Humidity	54% RH
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE

Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	TSC/22-23/7400-3	10-08-2023

ELECTRICAL SAFETY

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance >20MΩ	Measured values in MΩ 70	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 133	Uncertainty in % (±) 5.9
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 102	Uncertainty in % (±) 13.9

- REMARKS**
- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
 - 2.The user should determine the suitability of the instrument for its intended use.
 - 3.The recalibration interval should be determined on the user requirement.
 - 4.The results stated in this certificate relate only to the item calibrated.
 - 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer)
R.YAZINIYAN



Authorised Signatory
 Quality Manager
(D.VETRI SELVI)

Chief Executive