




CALIBRATION REPORT	
Report No : HBMS/CAL/22-23	Calibration date : 15.12.2022
	Calibration valid: 14.06.2023
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	<b>M/s, PRIMARY HEALTH CENTRE, S.S.KULAM, COIMBATORE.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
I Manufactured by	THERMO SCIENTIFIC
II Model	100-1000 µl
III Serial No	RW02769

**Details of the machine :** machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	1000.08
5	100.00	499.94	1000.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,

  
(Authorized signatory) 