

Date: 15/12/2022

To,

Apple Diagnostics Centre.

Ghatkopar, Mumbai.

Kind Attn: Dr. Amrish Mehta.

Calibration Certificate

We hear by certify that the Bio - Rad **D - 10** system **Sr. No. DC3E794719** Installed in your lab Is calibrated today as under & the performance was found satisfactory.

Flow calibration @ 1.5 ml / min (Flow checked by using volumetric variant sample cups with visible observation)

Thermo module temperature calibration @ 35°C (Temp. Checked by using D - 10 internal inbuilt temp. Sensor)

Detector log value @ 0.020 (Voltage verified by using Multimeter)

The next calibration is due on: 19/7/2023.

For,

Anand Chemiceutics

