



The Ultimate Calibration Services

The Quality Solutions....



CC-3020

Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.
Phone No.: 9890715627 / 9665003365, Mail : theucs20@gmail.com, info@theultimatecalibration.com • Website : www.theultimatecalibration.com
GST No. : 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/22060182 Date of Issue: 20-06-2022
Date of calibration: 16-06-2022 Suggested Due Date: 16-06-2023
ULR No:CC30202200000872F Location of calibration: Onsite

Name and Address of Customer: Dr. Amrish Mehta's Apple Diagnostic
2nd Floor, Prabhu Krupa Building, Tilak Road,
Opp. Silver Solitaire Building, Next to Bank of Baroda,
Ghatkopar (East), Mumbai-400 077.
Ref. No: Onsite
Date: 16-06-2022
Date of Receipt: 16-06-2022
Condition on Receipt: Good

Details of Equipment :

Name : Centrifuge

Make : Remi Model : R-8C Sr.No. : ZKN-32414
Id.No:--- Department : Clinical Pathology Location : ---
Least Count : 10 RPM Range : 0 to 6000 RPM Accuracy : Not Specified

Calibration Points: 1500, 2000, 2500, 3000, 3300 (As specified by customer)

Calibration Procedure Used: UCS/CAL/207

Details of STD equipment used :

Environmental Conditions : 25 ± 10 °C & 40 to 60% RH


Name	Id No	Parameter/Uncertainty	Certificate No	Valid up to
Digital Tachometer	UCS/CAL/047	Speed: ± 4.20 to 4.61 RPM	UCS/CAL/22020125	14-02-2023

The measurements are traceable to National/International standard through unbroken chain of traceability.


OBSERVATIONS :

Discipline : Mechanical		Parameter : Speed		Sub Group : Acceleration and Speed	
Range	Reading On UUC	Reading on Standard	Error	Expanded Uncertainty	
RPM	RPM	RPM	RPM	RPM	
0 to 6000	1500	1505	-5.0	± 7.21	
	2000	2005	-5.0	± 7.24	
	2500	2504	-4.0	± 7.29	
	3000	3008	-8.0	± 7.40	
	3300	3316	-16.0	± 7.55	

The Expanded uncertainty is calculated at the coverage factor k=2 at a confidence level of 95.45 %

Calibrated By: 
Name & Sign: Swapnil Malwankar



Authorised Signatory: 
Name & Designation: V.W. Patwardhan
Technical Manager

Format No: UCS/QF/7.8/01

Page no 1 of 2 P.T.O.



The Ultimate Calibration Services

The Quality Solutions....



CC-3020

Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.
Phone No.: 9890715627 / 9665003365, Mail : theucs20@gmail.com, info@theultimatecalibration.com • Website : www.theultimatecalibration.com
GST No. : 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/22060182

Date of Issue: 20-06-2022

Date of calibration: 16-06-2022

Suggested Due Date: 16-06-2023

ULR No:CC302022000000872F

Location of calibration: Onsite

Remarks:

1. The readings noted above are average of five readings.

The readings noted are particularly for the equipment submitted for calibration.


The reported values are valid at the time of calibration and under the controlled conditions.

Certificate should not be reproduced except in full without the prior approval of Laboratory.


Any handmade corrections made on the certificate, invalidates the Calibration Certificate.

The Ultimate Calibration Services will not be responsible for any mishandling of equipment after issuing the calibration certificate.

All precautions are being taken while handling the equipment at the time of calibration but lab is not liable for any misbehave of the equipment during calibration or any loss in the process of calibration.

Calibrated By : 
Name & Sign: Swapnil Malwankar



Authorised Signatory: 
Name & Designation: V.W.Patwardhan
Technical Manager

Format No: UCS/QF/7.8/01

-----End of Certificate-----

Page no 2 of 2