

The Ultimate Calibration Services

The Quality Solutions....



Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.

Phone No.: 9890715627 / 9665003365, Mail: theucs20@gmail.com, info@theultimatecalibration.com • Website: www.theultimatecalibration.com

GST No.: 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

CALIBRA	TION	CEDTIE	CATE
CALIBRA		CHRITH	ICAIH.

Calibration Certificate No: UCS/CAL/22060181

Date of calibration: 16-06-2022

ULR No:CC302022000000871F

Name and Address of Customer:

Dr. Amrish Mehta's Apple Diagnostic

2nd Floor, Prabhu Krupa Building, Tilak Road,

Opp. Silver Solitaire Building, Next to Bank of Baroda,

Ghatkopar (East), Mumbai-400 077.

Ref. No: Onsite

Date: 16-06-2022

Date of Receipt: 16-06-2022

Condition on Receipt: Good

Details of Equipment:

Name: Centrifuge

.

Make : Remi Model : R-8C Plus

Id.No:---

Department : Pathology

Location : ---

Least Count: 10 RPM

Range: 0 to 6000 RPM

Accuracy: Not Specified

Sr.No.: ZEAN-03612

Date of Issue: 20-06-2022

Suggested Due Date: 16-06-2023 Location of calibration: Onsite

Calibration Points: 1500, 2000, 2500, 3000, 3500 (As specified by customer)

Calibration Procedure Used: UCS/CAL/207

Details of STD equipment used:

Environmental Conditions: 25 ± 10 °C & 40 to 60% RH

Name	ld No	Parameter/Uncertainty	Certificate No	Valid up to
rigital Tachometer	UCS/CAL/047	Speed: ± 4.20 to 4.61 RPM	UCS/CAL/22020125	14-02-2023

The measurements are traceable to National/International standard through unbroken chain of traceability.

OBSERVATIONS:

Discipline : Med	chanical Parameter :	Parameter : Speed		Sub Group : Acceleration and Speed	
Range	Reading On UUC	eading On UUC Reading on Standard		Expanded Uncertainty	
RPM	RPM	RPM	RPM	RPM	
0 to 6000	1500	1504	-4.0	± 7.12	
	2000	2005	-5.0	± 7.34	
	2500	2506	-6.0	± 7.38	
	3000	3009	-9.0	± 7.71	
	3500	3507	-7.0	± 7.89	

The Expanded uncertainty is calculated at the coverage factor k=2 at a confidence level of 95.45 %

Calibrated By : မြေ<u>ပြယ်ရ</u>ဂုန်ငှ

Name & Sign: Swapnil Malwankar

Authorised Signatory:

Name & Designation: V.W.Patwa

Technical Manager

Format No: UCS/QF/7.8/01

Page no

1 of 2 P.T.O.



The Ultimate Calibration Services

The Quality Solutions....



Add.: Gala No. 1&2, Kamalkunj CHS, Opp. Union Bank of India, Near Kaka Gole Foundation, Kulgaon, Badlapur (E), Dist. Thane 421 503.

Phone No.: 9890715627 / 9665003365, Mail: theucs20@gmail.com, info@theultimatecalibration.com • Website: www.theultimatecalibration.com

GST No.: 27AANFT7678Q1ZT | PAN No.: AANFT7678Q

CALIBRATION CERTIFICATE

Calibration Certificate No: UCS/CAL/22060181

Date of calibration: 16-06-2022

ULR No:CC302022000000871F

Date of Issue: 20-06-2022

Suggested Due Date: 16-06-2023

Location of calibration: Onsite

Remarks:

1. The readings noted above are average of five readings.

The readings noted are particularly for the equipment submitted for calibration.

The reported values are valid at the time of calibration and under the controlled conditions.

Certificate should not be reproduced except in full without the prior approval of Laboratory.

Any handmade corrections made on the certificate, invalidates the Calibration Certificate.

The Ultimate Calibration Services will not be responsible for any mishandling of equipment after issueing the calibration certificate.

All precautions are being taken while handling the equipment at the time of calibration but lab is not liable for any misbehave of the equipment during calibration or any loss in the process of calibration.

Calibrated By: Plalwankar Name & Sign: Swapnil Malwankar CALIBRATION SERLI

Authorised Signatory:

Name & Designation: V.W.Patwardhan

Technical Manager

Format No: UCS/QF/7.8/01

-----End of Certificate-----

Page no

2 of 2