

VIEW LAB MONTHLY SUMMARY

Lab Name **TELANGANA DIAGNOSTIC HUB**
 Lab No **15729**
 Month **November**
 Year **2022**
 Constituent Group **Chemistry I**

Details About Robust Analysis
Detail About Monthly Summary
Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result **Print**

Date of Result Entered : 18/11/2022

Date of Report Published : 06/12/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	Hexokinase	Any Analyser (Automation / Semi Automation)	157	320.64	6.06	19.44	318 mg/dL	-0.14	3.10
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	672	56.51	9.35	5.28	64 mg/dL	1.42	0.41
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Any Analyser (Automation / Semi Automation)	535	3.94	10.35	0.41	4.59 mg/dL	1.59	0.04
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser (Automation / Semi Automation)	648	1.84	18.01	0.33	2.02 mg/dL	0.54	0.03
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	741	5.69	7.80	0.44	5.6 g/dL	-0.20	0.03
6	ALBUMIN	BCG - colorimetric	Any Analyser (Automation / Semi Automation)	743	3.43	8.07	0.28	3.4 g/dL	-0.11	0.02
7	CALCIUM	Arsenazo III	Any Analyser (Automation / Semi Automation)	621	10.18	7.58	0.77	10.7 mg/dL	0.67	0.06
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	767	9.00	10.83	0.97	9.2 mg/dL	0.21	0.07
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	749	116.64	7.85	9.16	122 mg/dL	0.59	0.67
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	742	220.65	8.78	19.38	237 mg/dL	0.84	1.42
11	HDL	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	521	28.19	15.91	4.49	29 mg/dL	0.18	0.39
12	AST	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	667	155.52	11.05	17.18	154 U/L	-0.09	1.33
13	ALT	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	671	48.21	17.48	8.43	53 U/L	0.57	0.65
14	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	503	139.79	15.45	21.59	140 U/L	0.01	1.93
15	AMYLASE	CNPG3	Any Analyser (Automation / Semi Automation)	435	91.00	17.68	16.09	104 U/L	0.81	1.54

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Pregr-D
8/12/22

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Lab Name **TELANGANA DIAGNOSTIC HUB**
 Lab No **15729**
 Month **November**
 Year **2022**
 Constituent Group **Thyroid Hormones & Cortisol**

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 18/11/2022

Date of Report Published : 06/12/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants CV	SD	Year Value	SDI	U
1	TSH	CLIA- Chemiluminescence immunoassay / CMIA	Abbott Architect (i1000 / i2000)	95	6.10	6.90	0.42	6.15 uIU/mL	0.12	0.09
2	TOTAL T3	CLIA- Chemiluminescence immunoassay / CMIA	Abbott Architect (i1000 / i2000)	80	93.24	5.59	5.21	91 ng/dL	-0.43	1.17

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Page 10
8/11/22