



PC-1033

111th IAMM EQAS Microbiology: Bacteriology/ Serology

CMC MICRO EQAS

Department of Clinical Microbiology, Christian Medical College, Vellore-632004, Tamil Nadu
Email: egas@cmcvellore.ac.in, Twitter: @microeqas, Phone: +91-416-2282588



JULY 2022

111th EQAS EVALUATION REPORT

MEMBER ID:

M 1 2 0 7

JULY 2022 / SEROLOGY

Test method employed for detection C-reactive protein (CRP) at your lab: Latex Agglutination
Peer group (n) = 229

Please refer the attached evaluation format for details on the criteria for evaluation

Marks Obtained: 10/10 (100%)

Parameter	Your Result	Your Value (mg/L)	Intended Result	Robust Mean	Robust SD	Range (mg/L)	Uncertainty of Assigned value	Z & Z'	Max Marks	Your Score
SE1	CRP	Negative	Not reported			Not Applicable			2	2
SE2	CRP	Positive	Positive	31.0011	19.1529	6 to 192	1.6328	-1.0	2	2
SE3	CRP	Positive	Positive	20.3592	12.0561	6 to 192	1.0525	-0.7	2	2

JULY 2022 / SEROLOGY

Test method employed for detection Rheumatoid Factor (RF) at your lab: Latex Agglutination
Peer group (n) = 406

Please refer the attached evaluation format for details on the criteria for evaluation

Parameter	Your Result	Your Value (IU/mL)	Intended Result	Robust Mean	Robust SD	Range (IU/mL)	Uncertainty of Assigned value	Z & Z'	Max Marks	Your Score
SE1	RF	Negative	Not reported			Not Applicable			2	2
SE2	RF	Negative	Not reported			Not Applicable			2	2
SE3	RF	Positive	Positive	16					2	2

NOT EVALUATED*

* As >30% participants scored <70% on this exercise, it has been excluded from the final evaluation.

Disclaimer:

This is a confidential document and subject to the rules of confidentiality as described by the ISO 17043:2010 standard.

MEMBER ID:

M 1 2 0 7

SE1	SE2	SE3	Marks obtained	
4	4	2	10	100%
4	4	2	Maximum marks = 10	




Dr. Rani Diana Sahni
Scientific Co-ordinator

Report Dispatch Date: 15.11.2022



Dr. John A Jude Prakash
Quality Manager



Dr. V. Balaji
PT Co-ordinator

***** End of Report *****