



## Calibrator Calibration

<b>Instrument:</b> 5 Part Haematology Analyzer	<b>Model:</b> BC -6200	<b>Serial No.:</b> TW-03000936	<b>Date of Cal.</b> 13.07.2022			
<b>Mode: OV-WB</b>	<b>Calibrator Lot No. PLUS0722</b>		<b>Exp.Date:05.08.2022</b>			
<b>Parameter</b>	<b>WBC</b>	<b>RBC</b>	<b>HGB</b>	<b>MCV</b>	<b>PLT</b>	<b>WBC-D</b>
TARGET	9.05	4.59	13.6	92.5	255	9.05
1.	8.12	4.60	14.2	94.0	228	8.70
2.	8.19	4.64	14.2	93.4	222	8.95
3.	8.14	4.60	14.1	93.0	222	8.72
4.	7.83	4.63	14.1	93.1	231	8.76
5.	7.90	4.63	14.2	92.9	215	8.48
6.	8.06	4.59	14.1	92.8	218	8.45
Mean	8.04	4.61	14.1	93.2	223	8.68
CV( %)	1.785	0.411	0.29	0.47	2.7	2.146
New Cal. Factor (%)	112.56	99.50	96.13	99.25	114.50	
Old Cal. Factors (%)	120.00	96.71	94.75	100.25	101.89	
Old TC1 (%)						101.40
New TC1 (%)						101.92

Telangana Diagnostic Hub JAGITYAL

For Rapid Diagnostic Pvt. Ltd.

G. Nagaraju  
Service Engineer



Rajaram Krishnamurthy  
National Service Manager



## Calibration Certificate

<b>Customer Name</b>	<b>Telangana Diagnostic Hub JAGITYAL</b>
Address	Near MCH Hospital, JAGITYAL-505327
Contact Person	Miss.S. Nivya
Contact No.	7989926226
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-03000936
Instrument Status	Under Warranty
Calibration Done On	13.07.2022
Next Calibration Due On	12.07.2023

**For Rapid Diagnostic Pvt. Ltd.**

*G. Nagaraju*

G. Nagaraju  
Service Engineer



*Rajaram Krishnamurthy*  
Rajaram Krishnamurthy  
National Service Manager