




# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com CC-3427



## Calibration Certificate

|                                    |  |                                |                        |                   |
|------------------------------------|--|--------------------------------|------------------------|-------------------|
| <b>CERTIFICATE NO.</b>             | : 221001A/N-105  | <b>Page</b>                    | : 1 of 1               |                   |
| <b>CALIBRATED ON.</b>              | : 02 / 10 / 2022   | <b>Ref. Std.</b>               | : IS:12508:1988        |                   |
| <b>CALIBRATION DUE.</b>            | : 01 / 10 / 2023   | <b>C.P. No.</b>                | : ACS/CP/10            |                   |
| <b>SRF No. &amp; DATE</b>          | : 221001A/N Dt. 01.10.2022   | <b>Phy. Inspection</b>         | : O.K.                 |                   |
| <b>ULR No. &amp; ISSUE DATE</b>    | : CC342722000000605F Dt. 14.10.2022  | <b>Cal. At</b>                 | : Site / Lab           |                   |
| <b>CALIBRATED FOR</b>              | : M/s Redcliffe Lifetech Pvt. Ltd.<br>Building No. 168, First Floor Sarathy,<br>9 Main Sector 6 Hrs Layout, Bangalore-560102 |                                |                        |                   |
| <b>ITEM CALIBRATED</b>             | : Centrifuge   |                                |                        |                   |
|                                    | <b>Make</b>  | : Remi                         | <b>Range</b>           | : (0 to 6000) RPM |
|                                    | <b>L. Count</b>  | :                              | <b>ID No.</b>          | : RCL/CP/Cent-02  |
|                                    | <b>Location</b>  | : Clinical Pathology           | <b>S. No.</b>          | : ZHHN-33045      |
| <b>ENVIRONMENT CONDITION:</b>      | Temp. { 25°C ± 3°C} R.H. {40% to 60%}  |                                |                        |                   |
| <b>STANDARD INSTRUMENTS USED :</b> |  |                                |                        |                   |
| <b>INSTT. NAME</b>                 | <b>ID No. / S. No.</b>   | <b>TRACEABILITY</b>            | <b>CERTIFICATE No.</b> | <b>DUE DATE</b>   |
| Tachometer                         | ACS/DTM/01/ 211-0155V  | Godrej                         | C-220912-6-1           | 11/09/2023        |
| <b>RESULTS</b>                     | : Mechanical Discipline  |                                |                        |                   |
|                                    | <b>* U. C. INSTT. READS (RPM)</b>  | <b>Standard Measured (RPM)</b> | <b>Error (RPM)</b>     |                   |
|                                    | 1000   | 1001.72                        | - 1.72                 |                   |
|                                    | 2000   | 2002.50                        | - 2.50                 |                   |
|                                    | 3000   | 3003.46                        | - 3.46                 |                   |
|                                    | 4000   | 4004.32                        | - 4.32                 |                   |
|                                    | (average of readings)  |                                |                        |                   |

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 3.09 rpm of reading at coverage Factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmary statement not provided since not required.

Prepared by

...End of Calibration Certificate....

