



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com CC-3427



Calibration Certificate

CERTIFICATE NO.	: 221001A/N-113	Page	: 1 of 1
CALIBRATED ON.	: 02 / 10 / 2022	Ref. Std.	: CG-11
CALIBRATION DUE.	: 01 / 10 / 2023	C.P. No.	: ACS/CP/07
SRF No. & DATE	: 221001A/N Dt. 01.10.2022	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342722000000613F Dt. 14.10.2022	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Building No. 168, First Floor Sarathy, 9 Main Sector 6 Hrs Layout, Bangalore-560102		
ITEM CALIBRATED	: Digital Temp. Indicator		
	Make : RoHS	Range	: (-50 to 110) °C
	L. Count :	ID No.	: RCL/SR/TEMP-06
	Location : Sample Receiving Area	S. No.	: N.A
ENVIRONMENT CONDITION:	Temp { 25°C + 3°C} R.H. {40% to 60%}		
STANDARD INSTRUMENTS USED :			
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2021/0784
			DUE DATE
			11/11/2022
RESULTS	: ET Discipline		
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)	
2.02	2	0.02	
4.04	4	0.04	
6.08	6	0.08	
8.12	8	0.12	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is $\pm 1.00^\circ\text{C}$ at coverage Factor $k=2$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

