

VIEW LAB MONTHLY SUMMARY

Lab Name	TELANGANA DIAGNOSTIC HUB		Details About Robust Analysis							
Lab No	14904		Detail About Monthly Summary							
Month	September		Detail about SDI							
Year	2022									
Constituent Group	Chemistry I									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 12/09/2022							
			Date of Report Published : 05/10/2022							
Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE	Hexokinase	Any Analyser (Automation / Semi Automation)	167	180.37	5.20	9.37	165 mg/dL	-1.64	1.45
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	661	98.36	8.92	8.77	100.95 mg/dL	0.30	0.68
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Any Analyser (Automation / Semi Automation)	561	3.11	10.12	0.32	2.81 mg/dL	-0.95	0.03
4	T.BILIRUBIN	Others (DPD, Vanadate Oxidation)	Any Analyser (Automation / Semi Automation)	231	3.08	12.33	0.38	3.4 mg/dL	0.84	0.05
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	805	5.20	9.33	0.48	4.2 g/dL	-2.06	0.03
6	ALBUMIN	BCG - colorimetric	Any Analyser (Automation / Semi Automation)	782	3.16	7.60	0.24	2.9 g/dL	-1.08	0.02
7	CALCIUM	OCPC (O - Cresolphthalein Compleazone)	Any Analyser (Automation / Semi Automation)	242	10.02	8.98	0.90	9.9 mg/dL	-0.13	0.12
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	739	6.76	11.49	0.78	6.1 mg/dL	-0.85	0.06
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	762	104.59	9.22	9.64	104 mg/dL	-0.06	0.70
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	735	177.25	9.98	17.68	176 mg/dL	-0.07	1.30
11	HDL	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	520	26.66	17.09	4.56	21 mg/dL	-1.24	0.40
12	AST	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	705	63.66	10.94	6.97	74 U/L	1.48	0.52
13	ALT	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	672	87.96	15.64	13.76	105 U/L	1.24	1.06

14	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	473	68.16	16.15	11.01	65 U/L	-0.29	1.01
15	AMYLASE	Enzymatic Colorimetric / G7PNP Blocked	Any Analyser (Automation / Semi Automation)	225	42.86	18.74	8.03	52 U/L	1.14	1.07
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

VIEW LAB MONTHLY SUMMARY

Lab Name	TELANGANA DIAGNOSTIC HUB		Details About Robust Analysis							
Lab No	14904		Detail About Monthly Summary							
Month	September		Detail about SDI							
Year	2022									
Constituent Group	HbA1c									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 12/09/2022							
			Date of Report Published : 05/10/2022							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	HbA1c	Enzymatic	Any Analyser	35	5.80	8.16	0.47	5.8 %	0.00	0.16
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out