



## Calibrator Calibration

<b>Instrument:</b> 5 Part Haematology Analyzer		<b>Model:</b> BC -6200		<b>Serial No.:</b> TW-04001042		<b>Date of Cal.</b> 13.07.2022
<b>Mode: OV-WB</b>		<b>Calibrator Lot No. PLUS0722</b>			<b>Exp.Date:05.08.2022</b>	
<b>Parameter</b>	<b>WBC</b>	<b>RBC</b>	<b>HGB</b>	<b>MCV</b>	<b>PLT</b>	<b>WBC-D</b>
<b>TARGET</b>	9.05	4.59	13.6	92.5	255	9.05
1.	8.92	4.58	14.2	94.3	226	7.96
2.	9.01	4.56	14.2	94.2	227	8.29
3.	8.82	4.54	14.1	94.0	226	8.28
4.	8.98	4.56	14.0	94.2	228	8.59
5.	8.84	4.60	14.1	94.2	219	8.36
6.	8.96	4.56	14.2	94.2	227	8.47
<b>Mean</b>	8.92	4.57	14.1	94.2	225	8.30
<b>CV( %)</b>	0.864	0.475	0.66	0.10	1.5	2.299
<b>New Cal. Factor (%)</b>	101.13	100.51	96.15	98.22	113.18	
<b>Old Cal. Factors (%)</b>	100.13	101.12	96.21	98.13	112.73	
<b>Old TC1 (%)</b>						102.77
<b>New TC1 (%)</b>						98.64

**For Rapid Diagnostic Pvt. Ltd.**

  
**N. Ramakrishna Reddy**  
 Service Engineer

  
**N. Mahender Reddy**  
 Zonal Service Manager



## Calibration Certificate

Customer Name	Telangana Diagnostic Hub Siddipet
Address	Area Hospital, SIDDIPET-502103
Contact Person	Mr. Srinivas Reddy
Contact No.	9246923882
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-04001042
Instrument Status	Under Warranty
Calibration Done On	13.07.2022
Next Calibration Due On	12.07.2023

**For Rapid Diagnostic Pvt. Ltd.**

  
N. Ramakrishna Reddy  
Service Engineer

  
N. Mahender Reddy  
Zonal Service Manager