



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 1576

Distribution No.: 157-D

Month/Year: August/2022

Instrument ID: MINDRAY BC 6200

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 15-10-2022[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	3.02	2.95	5.97	6.13	0.0170	-0.36	0.07	0.07	0.0050	0.00
RBC x10 ⁶ /µl	1	4.32	4.21	8.53	9.36	0.0080	-3.86	0.11	0.04	0.0020	1.89
Hb g/dl	1	11.2	11.2	22.4	22.6	0.0200	-0.39	0	0.1	0.0070	-1.35
HCT%	1	36.9	36.3	73.2	78.1	0.1120	-1.61	0.6	0.3	0.0210	1.01
MCV-fl	1	86.3	85.6	171.9	166.6	0.2100	0.88	0.7	0.3	0.0200	1.08
MCH-Pg	1	26.6	25.9	52.5	48.2	0.0450	3.63	0.7	0.2	0.0110	3.37
MCHC-g/dl	1	30.8	30.3	61.1	57.7	0.0940	1.27	0.5	0.3	0.0160	0.67
Plt. x10 ³ /µl	1	108	108	216	231	0.95	-0.59	0	5	0.28	-0.96
Retic %	2										

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=0 , Poly=27 L=10, E=2, Mono/Promono=2 , B1=15 P.M.=18, Mye=11, Meta=15, Other=	Blast: 43-80, Poly: 4-12, Lympho: 4-10, Promyelo: 0-12.25, Myelo: 1-6.5, nRBC/Mono/Meta/Eos: 0-5		
RBC Morphology	3	normocytic normpchromic, microcytes	Predominantly: Normocytic/Normochromic; Moderate: Microcytosis, Hypochromia; Mild: Anisocytosis, Macrocytosis		
Diagnosis	3	Acute Promyelocytic Leukemia	Acute Myeloid Leukemia (AML)		

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 157--D	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	354	349	82.81	92.26	3.72	3.44	13.47	4.3
RBC x10⁶/µl	1	354	354	87.85	88.42	6.21	5.08	5.94	6.5
Hb g/dl	1	354	354	85.31	90.68	5.93	3.95	8.76	5.37
HCT%	1	354	349	89.11	91.12	6.3	4.58	4.59	4.3
MCV-fl	1	354	349	91.12	93.41	5.73	2.58	3.15	4.01
MCH-Pg	1	354	349	85.67	93.98	8.31	2.29	6.02	3.73
MCHC-g/dl	1	354	349	89.68	91.4	5.44	2.87	4.88	5.73
Plt. x10³/µl	1	354	349	88.54	88.83	6.02	6.88	5.44	4.29
ReticCount%	2	354	232	89.22	95.26	9.05	9.91	1.73	-5.17
PS Assessment	3	354	338	Satisfactory :94.64%, Borderline Sat. :3.95%, Unsatisfactory :1.41%					

***Comments:**

1). Among Lab (EQA) : Results acceptable.

2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ± 2 : Acceptable, Z score ± 2 to ± 3 :Warning Signal, Z score $> \pm 3$: Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between " 0 to ± 2 " are texted in green colour. Z score value between " ± 2 to ± 3 " are texted in orange colour. Z score value $> \pm 3$ are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ($0.3*SDPA$). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value ($0.3*SDPA$).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----