




Calibrator Calibration

Instrument: 5 Part Haematology Analyzer		Model: BC -6200		Serial No.: TW-02000912		Date of Cal. 13.07.2022
Mode: OV-WB		Calibrator Lot No. PLUS0722			Exp.Date:05.08.2022	
Parameter	WBC	RBC	HGB	MCV	PLT	WBC-D
TARGET	9.05	4.59	13.6	92.5	255	9.05
1.	8.83	4.64	14.2	93.9	229	8.70
2.	8.91	4.68	14.4	93.9	230	8.95
3.	8.97	4.69	14.4	93.8	234	8.72
4.	8.86	4.67	14.3	93.9	223	8.76
5.	8.77	4.75	14.2	94.0	239	8.48
6.	8.75	4.68	14.3	93.8	227	8.45
Mean	8.85	4.69	14.3	93.9	230	8.68
CV(%)	0.944	0.758	0.4	0.08	2.4	2.146
New Cal. Factor (%)	102.28	97.94	95.14	98.52	110.81	
Old Cal. Factors (%)	101.50	88.00	95.00	99.50	98.96	
Old TC1 (%)						104.16
New TC1 (%)						101.96

For Rapid Diagnostic Pvt. Ltd.



G. Nagaraju
Service Engineer



N. Mahender Reddy
Zonal Service Manager




Calibration Certificate

Customer Name	Telangana Diagnostic Hub Sangareddy
Address	Govt. Hospital, SANGAREDDY-502001
Contact Person	Mr. Ramu
Contact No.	8074516990
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-02000912
Instrument Status	Under Warranty
Calibration Done On	13.07.2022
Next Calibration Due On	12.07.2023

For Rapid Diagnostic Pvt. Ltd.



G. Nagaraju
Service Engineer



N. Mahender Reddy
Zonal Service Manager