



## Calibrator Calibration

<b>Instrument:</b> 5 Part Haematology Analyzer	<b>Model:</b> BC -6200	<b>Serial No.:</b> TW-9A000724	<b>Date of Cal.</b> 13.07.2022			
<b>Mode: OV-WB</b>	<b>Calibrator Lot No. PLUS0722</b>			<b>Exp.Date:05.08.2022</b>		
<b>Parameter</b>	<b>WBC</b>	<b>RBC</b>	<b>HGB</b>	<b>MCV</b>	<b>PLT</b>	<b>WBC-D</b>
<b>TARGET</b>	9.05	4.59	13.6	92.5	255	9.05
1.	8.72	4.78	14.3	93.0	220	9.02
2.	8.51	4.66	14.3	91.4	220	8.62
3.	8.42	4.76	14.3	92.3	232	8.72
4.	8.74	4.72	14.3	90.8	215	8.86
5.	8.81	4.71	14.4	90.7	220	8.64
6.	9.01	4.68	14.3	91.0	221	9.04
<b>Mean</b>	8.70	4.72	14.3	91.5	222	8.82
<b>CV( %)</b>	2.436	0.998	0.26	0.99	2.5	2.099
<b>New Cal. Factor (%)</b>	104.00	97.25	94.92	101.05	115.12	
<b>Old Cal. Factors (%)</b>	103.92	92.23	93.21	99.82	98.84	
<b>Old TC1 (%)</b>						101.96
<b>New TC1 (%)</b>						98.70

For Rapid Diagnostic Pvt. Ltd.

G. Nagaraju  
Service Engineer

N. Mahender Reddy  
Zonal Service Manager

## Calibration Certificate



<b>Customer Name</b>	<b>Telangana Diagnostic Hub Karimnagar</b>
Address	Area Hospital, Karimnagar-505001
Contact Person	Sameera Nousheen
Contact No.	915422970
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-9A000724
Instrument Status	Under Warranty
Calibration Done On	13.07.2022
Next Calibration Due On	12.07.2023

**For Rapid Diagnostic Pvt. Ltd.**

G. Nagaraju  
Service Engineer

N. Mahender Reddy  
Zonal Service Manager