

CALIBRATION CERTIFICATE

DATE: 23.12.2022

NAME OF CUSTOMER:

SHREE SANJEVE CLINICAL LABORATORY,
 NO: 63/2, VIJAY TOWERS, 1ST FLOOR,
 PATTUKKOTTAI ROAD,
 ARANTHANGI.

BTS-350 CHEMISTRY SEMI AUTO ANALYZER PHOTOMETER BEARING.

S/NO:801759191CALIBRATION ON: 23.12.2022

CALIBRATION VALUES:

LIGHT INTENSITY TEST

LED	TINT	MAIN	REFERENCE
340	14	910496	181154
405	15	980515	695455
505	16	945952	551445
535	12	919952	451145
560	13	967841	662145
600	14	961561	512262
635	12	938452	551525
670	16	988511	581525

PERISTALTIC PUMP ADJUSTMENT TEST

VALUES	CURRENT	NEW
<u>SAMPLE VOLUME</u>	<u>44225</u>	<u>44225</u>
<u>POSITIONING</u>	<u>1502</u>	<u>1802</u>
<u>PUMB DELAY</u>	<u>2</u>	<u>2</u>

NEXT CALIBRATION DUE ON: 22.12.2023

TESTED BY: G DINESH



NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.