



CC-2725



CERTIFICATE OF CALIBRATION

Certificate No: CC-22-23-S-1218/03

Page. No : 1 of 1

Customer Name & Address M/s. SHREE SANJEVE CLINICAL LABORATORY No: 63/2, Vijay Towers, First Floor, Pattukkottai Road, Aranthangi - 614 616.	ULR No.	CC272522000011084F
	SRF No.	VCI-22-23-S-1218
	Date of Receipt	23.12.2022
	Date of Calibration	23.12.2022
	Due Date of Calibration	22.12.2023
	Date of Issue	24.12.2022

Details of Unit Under Calibration

Description	Dry Bath Incubator	Make	Lab Line
Range	Ambient to 100°C	Model	DBI-22
Resolution	0.1°C	Condition of UUC	Good
Serial Number	5065	Instrument Location	--
ID Number	--	Cal. At	On Site

Environmental Condition		Calibration Method Used	
Temperature (°C)	30±5	National / International Standard	DKD-R-5-7
Humidity (%RH)	30 - 75	Cal Procedure No	VCI-L03-WI-THER-005

Standard Used

Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	RTD Sensor with Indicator	VCI-T-RTD-002& VCI-ET-DT-013	CC-22-23-L-015/03	Aparajit Instruments	Venus Calibration CC-2725	08.01.2023

Results of Calibration

Sl. No.	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Measurement Uncertainty ± (°C)
01	30.12	30.0	-0.12	0.36
02	37.26	37.0	-0.26	0.36
03	50.39	50.0	-0.39	0.36
04	80.46	80.0	-0.46	0.36
05	100.53	100.0	-0.53	0.36

Remarks:

- UUC is Defined as the Unit Under Calibration.
- This Calibration Certificate Results relates only to the above UUC & This report shall not be reproduced except in full/ part without approval of the Venus Calibration and Instruments.
- The reported results are valid at the time of under stated condition of measurement.
- Calibration of the UUC are traceable to National / International Standards.
- The recalibration interval shall be decided by the user.
- Nabl-133 guidelines are adopted for use of NABL symbol.
- The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor $k=2$.
- Temperature Chart : ITS-90

Checked & Issued by:

R. Bhuvaneshwari
(Engineer - Calibration)

*****END OF THE CERTIFICATE*****



Authorised by:

C. Dhinakaran

(Technical / Quality Manager)

No.16, 6th Street, Annai Sathya Nagar, Poothapedu Ramapuram, Chennai - 89.

Ph: +91 44 2476 0533 /0633, Mob: 90031 20649, 9789893643, 8807402558

e: venuscallab@gmail.com. info@venuscal.com | www.venuscal.com

Scanned by TapScanner