



CC-2725

**CERTIFICATE OF CALIBRATION**

Certificate No: CC-22-23-S-1218/01

Page. No : 1 of 1

Customer Name & Address M/S. SHREE SANJEVE CLINICAL LABORATORY No: 63/2, Vijay Towers, First Floor, Pattukkottai Road, Aranthangi - 614 616.	ULR No.	CC272522000011082F
	SRF No.	VCI-22-23-S-1218
	Date of Receipt	23.12.2022
	Date of Calibration	23.12.2022
	Due Date of Calibration	22.12.2023
	Date of Issue	24.12.2022

Details of Unit Under Calibration

Description	Refrigerator	Make	LG
Range	2 to 8°C	Model	--
Resolution	NA	Condition of UUC	Good
Serial Number	--	Instrument Location	--
ID Number	RF-01	Cal. At	On Site

Environmental Condition		Calibration Method Used	
Temperature (°C)	30±5	National / International Standard	DKD-R-5-7
Humidity (%RH)	30 - 75	Cal Procedure No	VCI-L03-WI-THER-005

Standard Used

Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	RTD Sensor with Indicator	VCI-T-RTD-002& VCI-ET-DT-013	CC-22-23-L-015/03	Aparajit Instruments	Venus Calibration CC-2725	08.01.2023

Result of Calibration

S. No	UUC Setting (°C)	Standard Reading in °C	Measurement Uncertainty ± (°C)
1	2 to 8	4.42	0.36

Remarks:

- UUC is Defined as the Unit Under Calibration.
- This Calibration Certificate Results relates only to the above UUC & This report shall not be reproduced except in full/ part without approval of the Venus Calibration and Instruments.
- The reported results are valid at the time of under stated condition of measurement.
- Calibration of the UUC are traceable to National / International Standards.
- The recalibration interval shall be decided by the user.
- Nabl-133 guidelines are adopted for use of NABL symbol.
- The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor $k=2$.
- Temperature Chart : ITS-90

Checked & Issued by:


 R. Bhuvaneshwari
 (Engineer - Calibration)

*****END OF THE CERTIFICATE*****



Authorised by:


 C. Dhinakaran

(Technical / Quality Manager)

No.16, 6th Street, Annai Sathya Nagar, Poothapedu Ramapuram, Chennai - 89.

Ph: +91 44 2476 0533 /0633, Mob: 90031 20649, 9789893643, 8807402558

e: venuscallab@gmail.com. info@venuscal.com | www.venuscal.com

Scanned by TapScanner