



CERTIFICATE OF CALIBRATION

| Certificate No: CC-22-23-S-1218/02 | | Page. No: 1 of 1 | |
|---|-------------------------|--------------------|--|
| Customer Name & Address M/s. SHREE SANJEVE CLINICAL LABORATORY No: 63/2, Vijay Towers, First Floor, Pattukkottai Road, Aranthangi - 614 616. | ULR No. | CC272522000011083F | |
| | SRF No. | VCI-22-23-S-1218 | |
| | Date of Receipt | 23.12.2022 | |
| | Date of Calibration | 23.12.2022 | |
| | Due Date of Calibration | 22.12.2023 | |
| | Date of Issue | 24.12.2022 | |

| Description | Refrigirator Centrifuge | Make | Chem-Leon | |
|---------------|-------------------------|---------------------|-----------|--|
| Range | 0 to 6000 rpm | Model | A8 | |
| Resolution | 100 rpm | Condition of UUC | Good | |
| Serial Number | 2224V29 | Instrument Location | | |
| ID Number | ** | Cal. At | On Site | |

| Environmental Condition | | Calibration Method Used | | |
|-------------------------|---------|-----------------------------------|--------------------|--|
| Temperature (°C) | 23±5 | National / International Standard | SANAS TR 45-02 | |
| Humidity (%RH) | 40 - 70 | Cal Procedure No | VCI-L03-WI-SPE-002 | |

Standard Used

| Sl. No. | Description | ID.No. / SI. No. | Certificate No. | Make/Model | Traceability | Valid till |
|---------|--------------------|------------------|-------------------|-------------------|-------------------------------|------------|
| 1 | Digital Tachometer | VCI-M-TM-011 | CC-22-23-L-015/08 | Lutron / DT-2234C | Venus Calibration CC- 2725 | 08.01.2023 |

| Results of Calibration | | | | |
|------------------------|-------------------|------------------------|----------------|---------------------------------|
| Sl. No. | UUC Reading (rpm) | Standard Reading (rpm) | Error (rpm) | Measurement Uncertainty (±) (%) |
| 01 | 1000 | 1001 | -1 | 5.78 |
| 02 | 2000 | 2001 | -1 | 2.90 |
| 03 | 3000 | 3002 | -2 | 1.97 |
| 04 | 5000 | 5003 | -3 | 1.25 |
| 05 | 6000 | 6004 | -4 | 1.14 |

Remarks:

- 1. UUC is Defined as the Unit Under Calibration.
- This Calibration Certificate Results relates only to the above UUC & This report shall not be reproduced except in full/ part without approval of the Venus Calibration and Instruments.
- 3. The reported results are valid at the time of under stated condition of measurement.
- 4. Calibration of the UUC are traceable to National / International Standards.
- 5. The recalibration interval shall be decided by the user.
- Nabl-133 guidelines are adopted for use of NABL symbol.
- 7. The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor k = 2.

Checked & Issued by:

R.Bhuvaneshwari (Engineer - Calibration)

********END OF THE CERTIFICATE****

2476 0533 TON & SAN TON BURN THE SAN T

Authorised by:

C.Dhinakaran (Technical / Quality Manager)

No.16, 6th Street, Annai Sathya Nagar, Poothapedu Ramapuram, Chennai - 89.
Ph: +91 44 2476 0533 /0633, Mob: 90031 20649, 9789893643, 8807402558
e: venuscallab@gmail.com_info@venuscal.com_l_www.venuscal.c