



## CERTIFICATE OF CALIBRATION/VERIFICATION

Customer Name: Inlaks General Hospital, Chembur  
Model: Fully Automated Bio-Chemistry Analyzer-EM 200  
Serial No: S2001038  
Calibration Date: 16-11-2022  
Next Calibration due on before : 15-11-2023  
Calibrator Used: Multical Lot Number: 2204083  
Quality Control Used: Lot Number:S062125B/S062131B

This is to certify that the above mentioned instrument has been verified of calibration according to the procedures provided by the manufacturer -

### General Parameters:

Parameter	Measured Value	Acceptable Range	
Line Voltage	230V	230 V +/- 5%	Pass
Logic Voltage	5.0 V	4.9 V to 5.1 V	Pass
Lamp Voltage	12.0 V	11.8 V to 12.2 V	Pass
Flow cell Temperature	37.0 deg	37 +/- 0.5	Pass
45 Cuvette blank value for for all filter	Enclose value sheet for all cuvette	Abs 0.02-0.2	Pass