

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Vinayaga Micro Clinical Lab, Attur Main Road, Mangalapuram	Certificate No:	AMS/HOSP/289/1993
		Issuing Date:	03/09/2022
		Page No:	1 of 2

Details of Device under Calibration (DUC):

DUC:	Bio Chemistry Analyser	Id. No:	VMC/LAB/BCA/001
Make:	VChem NExi	Department:	Laboratory
Model:	Vector	Calibrated On:	01/09/2022
Sl. No:	ES1052107CN1119	Calibration Due Suggested:	31/08/2023
Type:	Semi Auto	DUC Condition on Receipt:	OK
Range & Least Count:	---		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/6908/2022	31/01/2023
2	Manometer	Lutron/PM9100	PC/01/22/ATL/0432-05	30/01/2023
3	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
4	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	278	mA	500	Normal
2	Frequency	49.9	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	236.1	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	235.5	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	4.9	Volts	< 7	Normal

*Electrical Safety: Please check the attached log sheet

SL.No	Parameter	Measured	Standard	Correction	Unit
1	Lamp Voltage	11.78	12.0	-0.22	Volts
2	Detector Sensitivity	0.0	0.0	0.0	% Absorbance
3	Sipper Volume	432.5	450	-17.5	µl
4	Peristaltic Pump pressure	-14.42	-15.0	-0.58	cmH2O

Allowable Tolerance: set by Customer

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

aceagm@gmail.com



www.acemedical.in

Approved by:
(Technical Manager)

93455 56140, 93455 56148