



Calibrator Calibration

Instrument: 5Part Haematology Analyzer		Model: BC-6200		SerialNo.: TW-03000935		Date of Cal. 27.12.2022
Mode:OV-WB		Calibrator Lot No. PLUS1222			Exp. Date: 05.01.2023	
Parameter	WBC	RBC	HGB	MCV	PLT	WBC-D
TARGET	9.05	4.59	13.6	92.5	255	9.05
1.	8.12	4.60	14.2	94.0	228	8.70
2.	8.19	4.64	14.2	93.4	222	8.95
3.	8.14	4.60	14.1	93.0	222	8.72
4.	7.83	4.63	14.1	93.1	231	8.76
5.	7.90	4.63	14.2	92.9	215	8.48
6.	8.06	4.59	14.1	92.8	218	8.45
Mean	8.04	4.61	14.1	93.2	223	8.68
CV(%)	1.785	0.411	0.29	0.47	2.7	2.146
New Cal. Factor (%)	112.56	99.50	96.13	99.25	114.50	
Old Cal. Factors (%)	120.00	96.71	94.75	100.25	101.89	
Old TC1 (%)						101.40
New TC1 (%)						101.92

For Rapid Diagnostic Pvt. Ltd.

G. Nagaraju
Service Engineer

N. Mahender Reddy
Zonal Service Manager



Calibration Certificate

CustomerName	TelanganaDiagnostic Hub Gadwal
Address	GADWAL-509125
ContactPerson	G.RAVIKUMAR
ContactNo.	8179625241
Instrument	5PartHaematologyAnalyzer
Model	BC6200
SerialNo.	TW-03000935
InstrumentStatus	UnderWarranty
CalibrationDoneOn	27.12.2022
NextCalibrationDueOn	26.12.2023

ForRapidDiagnosticPvt. Ltd.

G.Nagaraju
ServiceEngineer

N.MahenderReddy
ZonalServiceManager