Ace Medical Solutions (India) Pvt. Ltd.

. An ISO 9001-2015 Certified Company

No: 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.





GSTIN. 33AAHCA8053H1ZR

DL No. MDU/4205/20B,MDU/3971/21B

CIN: U33110TN2009PTC072140

CALIBRATION CERTIFICATE

	M/S. Vinayaga Micro Clinical Lab,	Certificate No:	AMS/LAB/289/1995
Name &	Attur Main Road,	Issuing Date:	03/09/2022
Address:	Mangalapuram	Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Microscope	Id. No:	VMCL/LAB/MS/001
Make:	SH-Brand	Department:	Laboratory
Model:	Led	Calibrated On:	01/09/2022
Sl. No:		Calibration Due	02/07/2022
Type:	Binocular Opto Sys.	Suggested:	31/08/2023
Range &		DUC Condition	
Least Count:		on Receipt:	OK

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1 .	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/6908/2022	31/01/2023
2	Digital Lux Meter	HTC/ LX - 103	PC/01/22/ATL/0265-02	30/01/2023
3	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60%

(At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	238	mA	300	Normal
2	Frequency	50.0	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	231.5	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	238.4	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	8.8	Volts	< 7	Normal

^{*}Electrical Safety: Please check the attached log sheet

Test Results: Light Intensity LUX:

SL.No.	Knob Position	DUC Reading (LUX)	Distance (Cm)
1	Low	400	30
2	Med.	3600	-30
3	High	6200	A30

Note:

- 1. Calibration Points were selected as per customer request.
- 2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
- 3. Results Reported are valid at the time of and under the stated conditions of measurements,

Calibrated By:

(Calibration Engineer)

Approved by: (Technical Manager)

93455 56140, 93455 56148

aceagm@gmail.com