

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Vinayaga Micro Clinical Lab, Attur Main Road, Mangalapuram	Certificate No:	AMS/LAB/289/1992
		Issuing Date:	03/09/2022
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	VMC/LAB/VMP/003
Make:	----	Department:	Laboratory
Model:	----	Calibrated On:	01/09/2022
Sl. No:	051715407	Calibration Due Suggested:	31/08/2023
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	100 – 1000µl		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

### Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	100.0	110.00	110.00	+10.00
2	200.0	210.00	210.00	+10.00
3	500.0	500.00	500.00	+00.00
4	800.0	800.00	800.00	+00.00
5	1000.0	990.00	990.00	-10.00

Allowable Tolerance: User Defined

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)

### Approved by:

(Technical Manager)

