DiaSys Diagnostics India Private Limited 229, FIA, Patparpanj Industrial Area. Ground and First Floor, New Delhi -110092



Calibration Certificate

Following voltages have been checked & adjusted on DiaSys SYS 200 Pro (Sl. No. SYS20220807) at Origin Path Lab, HB Twin Tower 2nd, 4rth floor, Netall Subhash palace, Delhi

Calibration Report:

System voltages are in the specified range - Checked by Digital Multimeter are as below:

S.No.	Module	Observed Value	Acceptable Range
1	Power supply board o/p	+24.1 VDC	+24.0 +/- 0.5 VDC
		+05.0 VDC	+05.0 +/- 0.2 VDC
2	Lamp voltage	+12.1 VDC	12.0 +/- 0.5 VDC

System Temperature values are in the specified range –Checked by Digital Thermometer are as below:

S.No.	Module	Observed Value	Acceptable Range
1	Reaction Disk Temp	37.0 C	37.0 C +/- 0.3 C

Photometer Adjustment Values are within range as recommended:

S.No.	Module	Observed Value	Acceptable Range
1	ADC Count for all Wavelengths	10,240 to 10,300	8000 to 18,000

- All other mechanisms have been checked through field service program & Software, found OK.
- > Performed calibration for End Point, Kinetic & Fixed Time Kinetic. Run control & sample. Results were satisfactory.
- Next Calibration will be performed on 01/09/2023.

For DiaSys Diagnostics India Pvt. Ltd.

Asmat Sami

Date: 02/09/2022

Regional Service Manager www.DiaSys-Diagnostic.com Contact: 022 3371 4323 Toll Free: 1800 120 1447

Registered Office

Plot No. A-821, T.T.C. Industrial Area, MIDC, Mahape, Navi Mumbai - 400710

Maharashtra, India

DiaSys Diagnostics India Private Limited

The Specialists for diagnostic system solutions Point-of care testing and automated Laboratory systems CIN

: U74999MH2013PTC248371

Email Tel. No. ; info@desys in

Toll Free www.diasys-diagnostics.com

:+91-022-71068331 : 1800 120 1447