



AVM LABS PVT. LTD.

No. 49, 3rd Street | Moorthy Nagar | Chettiar Agaram | Porur | Chennai - 600 077.
Mob.: +91 770 893 2211 | 735 837 6115
Email: avmlabsales@gmail.com | avmlabs21@gmail.com | www.avmgroup.biz



CC20882200028372F

CALIBRATION CERTIFICATE

CERT No.	: C-22/23-1189-02	DATE OF ISSUE	: 8-Dec-22	PAGE	: 1 OF 1
JOB No.	: J-22/23-1189	CALIBRATED ON	: 8-Dec-22	SUGG.NEXT CAL	: 8-Dec-23

DETAILS OF CUSTOMER :			
NAME & ADDRESS	: M/s. MEDITECH DIAGNOSTIC CENTRE SHILPA COMPLEX, NEAR PANJAYATH JUNCTION MUPPATHADAM - P.O. ERNAKULAM DISTRICT, ALUVA - 683 110.	CALIBRATED AT	: IN LAB
		DATE OF RECEIPT	: 8-Dec-22
		CONDITION OF RECEIPT	: SATISFACTORY
		CALIB. PROCEDURE	: AVM-WI-M13

DETAILS OF DEVICE UNDER CALIBRATION :			
NOMENCLATURE	: MICROPIPETTE	RANGE / SIZE	: 100 ~ 1000 µl
MANUFACTURER	: ERBA	OPERATING RANGE	: -
MODEL	: -	INPUT / OUTPUT	: -
TYPE	: -	RESOLUTION	: 1µl
SERIAL NUMBER	: YESF548376	ACCURACY	: -
ID/TAG NO.	: MED-004	LOCATION	: CLINICAL BIOCHEMISTRY

ENVIRONMENTAL CONDITIONS:			
TEMPERATURE :	22.8 °C	HUMIDITY :	55 %RH

MEASUREMENT TRACEABILITY			
The reference measurement standards used are traceable to National/International Standards.			
REFERENCE STANDARDS	SERIAL / TAG NUMBER	CERTIFICATE NUMBER	DUE DATE
ELECTRONIC SEMI MICRO BALANCE	327560/11 / BI-0094	C-AVM-IH-22/23-19	16-Jul-23

CALIBRATION RESULTS			
Nominal Capacity	Measured Capacity (Mean)	Error	Expanded Uncertainty (±)
µl	µl	µl	µl
100	99.29	-0.71	5.05
500	499.57	-0.43	1.16
1000	999.67	-0.33	2.68

The reported Expanded Uncertainty is calculated at 95.45% Confidence Level with a coverage factor of k = 2.

Note : The instrument is within the Permissible error as per the standard ISO 8655.

END OF CERTIFICATE

CALIBRATED BY: S. MANOJ CALIBRATION ENGINEER	REVIEWED & APPROVED BY: S.JAGATHEESWARI / AVICHI MEIYAPPAN AUTHORISED SIGNATORY
---	--