



AVM LABS PVT. LTD.

No. 49, 3rd Street | Moorthy Nagar | Chettiar Agaram | Porur | Chennai - 600 077.
Mob.: +91 770 893 2211 | 735 837 6115
Email : avmlabsales@gmail.com | avmlabs21@gmail.com | www.avmgroup.biz



C.G. 7-2088

CALIBRATION CERTIFICATE

ULR No.: CC208822000028378F

CERT No.	: C-22/23-1189-08	DATE OF ISSUE	: 8-Dec-22	PAGE	: 1 OF 1
JOB No.	: 1-22/23-1189	CALIBRATED ON	: 8-Dec-22	SUGG.NEXT CAL	: 8-Dec-23

DETAILS OF CUSTOMER :

NAME & ADDRESS	: M/s. MEDITECH DIAGNOSTIC CENTRE SHILPA COMPLEX, NEAR PANJAYATH JUNCTION MUPPATHADAM - P.O. ERNAKULAM DISTRICT, ALUVA - 683 110.	CALIBRATED AT	: IN LAB
		DATE OF RECEIPT	: 8-Dec-22
		CONDITION OF RECEIPT	: SATISFACTORY
		CALIB. PROCEDURE	: AVM-WI-T06

DETAILS OF DEVICE UNDER CALIBRATION :

NOMENCLATURE	: THERMO HYGROMETER	RANGE / SIZE	: -10 ~ 50 °C / 10 ~ 99 %RH
MANUFACTURER	: APTECDEALS	OPERATING RANGE	: -
MODEL	: HTC-1	INPUT / OUTPUT	: -
TYPE	: -	RESOLUTION	: 0.1°C / 1 %RH
SERIAL NUMBER	: -	ACCURACY	: -
ID/TAG NO.	: MED-010	LOCATION	: CLINICAL BIOCHEMISTRY

ENVIRONMENTAL CONDITIONS:

TEMPERATURE :	22.5 °C	HUMIDITY :	55 %RH
---------------	---------	------------	--------

MEASUREMENT TRACEABILITY

The reference measurement standards used are traceable to National/International Standards.

REFERENCE STANDARDS	TAG NUMBER	CERTIFICATE NUMBER	DUE DATE
HUMIDITY & TEMPERATURE METER WITH PROBE	EMS082K160939 / BL-0284	TSC/22-23/7913-12	19-Aug-23

CALIBRATION RESULTS

1. TEMPERATURE @ 50%RH

STD VALUE (REF)	TEST VALUE (DUC)(MEAN)	ERROR	EXPANDED UNCERTAINTY (±)
°C	°C	°C	°C
15.21	15.3	0.09	0.79
30.41	30.5	0.09	0.79
48.58	48.7	0.12	0.79

2. HUMIDITY @ 25°C

STD VALUE (REF)	TEST VALUE (DUC)(MEAN)	ERROR	EXPANDED UNCERTAINTY (±)
%RH	%RH	%RH	%RH
20.15	20	-0.15	1.63
60.32	60	-0.32	1.63
90.47	90	-0.47	1.63

The reported Expanded Uncertainty is calculated at 95.45% Confidence Level with a coverage factor of k = 2.

END OF CERTIFICATE

CALIBRATED BY:

M. SIVAMANKANDAN
CALIBRATION ENGINEER

REVIEWED & APPROVED BY :

S. JAGATHEESWARI / AVICHI MEIYAPPAN
AUTHORISED SIGNATORY