

Date: 17th July 2022

Calibration Certificate

This is to certify that Automated Hematology Analyzer **XN-550** Serial No: **18115** installed at **Jai Prakash District Hospital Bhopal** has been calibrated.

Calibration Done on 17th July 2022

Calibration Validity: 17th July 2022 to 16th July 2023

Calibration Due on 16th July 2023

WHOLE BLOOD CALIBRATION DETAIL

LOT NO.- 21642101 VALID UP TO- 17/07/2022 EXPIRY- 17/07/2022

Parameter	WBC	RBC	HGB	HCT	PLT
1	7.63	4.44	12.4	38.2	252
2	7.47	4.41	12.3	38.0	244
3	7.82	4.37	12.4	37.7	244
4	7.78	4.38	12.4	37.7	248
5	7.69	4.36	12.4	37.5	250
6	7.79	4.35	12.3	37.4	250
7	7.61	4.40	12.4	37.8	249
8	7.67	4.41	12.4	37.9	255
9	7.63	4.36	12.5	37.5	255
10	7.61	4.37	12.3	37.7	249

11	7.63	4.34	12.3	37.3	246
MEAN	7.670	4.375	12.37	37.65	249.0
TARGET	7.577	4.408	12.43	36.19	249.7
CV %	1.4	0.6	0.6	0.6	1.6
Acceptable Limit %	3.0	1.5	1.0	1.5	4.0

Calibration accepted for above mentioned values

For Sysmex India Pvt. Ltd.

Authorized Signatory



Sysmex India Pvt Ltd (CIN: U33120MH1998PTC115943)

Office 1002, Damji Shamji Business Galleria, 10th Floor, LBS Marg, Kanjurmarg (West), Mumbai – 400078, Maharashtra, India Tel. +91 (22) 6112 6666

Factory Village Malpur, Pargana Dharampur, Nalagarh Road, Baddi – 173205, Dist. Solan, Himachal Pradesh, India Tel. +91-9218422282 / 9816672282

www.sysmex.co.in
