

CAC

BIO-RAD

BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	District hospital Bhopal
Address:	1250 Link Road No. 2 Tulsi nagar
Province-Postal code	462001
Contact Name	Mr. Abhinav

Calibration Certificate

Date	01-09-2022 <small>(dd mmmm, yyyy)</small>	Instrument Model	D-10 Hb Analyzer	Instrument Serial Number	DM0G519494
------	--	------------------	------------------	--------------------------	------------

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
Glassco/measuring cylinder	11.20/2121	21/11/2020

Specified Range@ 1.5ml/Minute	Measured Value
1.425 – 1.575 ml/minute	1.482

Column Holder temperature Reference: Service Manual: Section 7.9

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
Fluke - 179	28400220	23-01-22

Specified Range @ 35°C	Measured Value
33.4°C – 36.4°C	34.9

Detector log value

Reference: Service Manual: Section 7.2 or
Field Bulletin D-10-081

Measured with	Specified Range	Measured Value
Instrument Service Software	0.018 – 0.022 AU	0.0200

The next calibration is due on	01-09-2023	<small>(dd mmmm, yyyy)</small>
--------------------------------	------------	--------------------------------

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory:

AB

01/09/22

ABHINAV PUROHIT
(Division Lab Co-ordinator-BHOPAL)

POCT SCIENCE HOUSE PVT. LTD.
MAN. NO. 8827251031

Calibration Certificate