

VIEW LAB MONTHLY SUMMARY

Lab Name	GOVT AREA HOSPITAL	Details About Robust Analysis									
Lab No	14947	Detail About Monthly Summary									
Month	November	Detail about SDI									
Year	2022										
Constituent Group	Chemistry I										
		Click on the analyte to view Graphical Data				All Analyser Result		Print			
		Date of Result Entered : 21/11/2022									
		Date of Report Published : 06/12/2022									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U	
1	GLUCOSE	Hexokinase	Any Analyser (Automation / Semi Automation)	157	320.64	6.06	19.44	317 mg/dL	-0.19	3.10	
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	672	56.51	9.35	5.28	53.5 mg/dL	-0.57	0.41	
3	CREATININE	Enzymatic Colorimetric	Any Analyser (Automation / Semi Automation)	122	4.04	6.14	0.25	3.68 mg/dL	-1.45	0.04	
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser (Automation / Semi Automation)	648	1.84	18.01	0.33	1.7 mg/dL	-0.42	0.03	
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	741	5.69	7.80	0.44	5.5 g/dL	-0.43	0.03	
6	ALBUMIN	BCG - colorimetric	Any Analyser (Automation / Semi Automation)	743	3.43	8.07	0.28	3.2 g/dL	-0.83	0.02	
7	CALCIUM	Arsenazo III	Any Analyser (Automation / Semi Automation)	621	10.18	7.58	0.77	10.2 mg/dL	0.03	0.06	
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	767	9.00	10.83	0.97	8.8 mg/dL	-0.21	0.07	
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	749	116.64	7.85	9.16	111 mg/dL	-0.62	0.67	
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	742	220.65	8.78	19.38	222 mg/dL	0.07	1.42	
11	HDL	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	521	28.19	15.91	4.49	30 mg/dL	0.40	0.39	
12	AST	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	667	155.52	11.05	17.18	145 U/L	-0.61	1.33	
13	ALT	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	671	48.21	17.48	8.43	47 U/L	-0.14	0.65	

14	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	503	139.79	15.45	21.59	126 U/L	-0.64	1.93
15	AMYLASE	CNPG3	Any Analyser (Automation / Semi Automation)	435	91.00	17.68	16.09	99 U/L	0.50	1.54
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out