



## Calibrator Calibration

<b>Instrument:</b> 5 Part Haematology Analyzer	<b>Model:</b> BC -6200	<b>Serial No.:</b> TW-03000934	<b>Date of Cal.</b> 09.07.2022			
<b>Mode: OV-WB</b>	<b>Calibrator Lot No. PLUS0722</b>			<b>Exp.Date:05.08.2022</b>		
Parameter	WBC	RBC	HGB	MCV	PLT	WBC-D
TARGET	9.05	4.59	13.6	92.5	255	9.05
1.	8.82	4.92	14.5	92.5	247	7.94
2.	8.89	4.81	14.5	92.7	259	8.26
3.	8.92	4.81	14.4	92.2	249	8.29
4.	8.65	4.86	14.4	92.6	259	8.49
5.	8.81	4.80	14.4	92.7	253	8.33
6.	8.63	4.85	14.4	92.3	253	8.47
Mean	8.79	4.84	14.4	92.5	253	8.30
CV( %)	1.378	0.923	0.31	0.21	1.9	2.372
New Cal. Factor (%)	103.00	94.77	94.31	99.97	100.63	
Old Cal. Factors (%)	103.41	94.38	95.29	100.05	99.55	
Old TC1 (%)						102.14
New TC1 (%)						105.92

For Rapid Diagnostic Pvt. Ltd.

G. Nagaraju  
Service Engineer

N. Mahender Reddy  
Zonal Service Manager



## Calibration Certificate

Customer Name	Telangana Diagnostic Hub Mulugu
Address	Area Hospital, Mulugu-506343
Contact Person	Mahathi Mothe
Contact No.	8328482979
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-03000934
Instrument Status	Under Warranty
Calibration Done On	09.07.2022
Next Calibration Due On	08.07.2023

For Rapid Diagnostic Pvt. Ltd.



G. Nagaraju  
Service Engineer



N. Mahender Reddy  
Zonal Service Manager