



CERTIFICATE OF CALIBRATION

Issued To/Hospital Name: ICTC LAB DIST. HQRS. HOSPITAL, KARIMNAGAR.

Certificate No	FBM/C22/4324	Equipment Name	CENTRIFUGE
Serial Number	1072	Model	NA
Manufacturer	BAROMETER	Department	LAB

Calibration Standards used:

S.No	Test Equipment	Manufacturer	Model No	Serial No	Cal Done Date	Cal Due Date
1	DIGITAL TACHOMETER	AMPROBE	TACH-10	63215,	23.06.2022	23.06.2023

"PERFORMANCE OF THE EQUIPMENT FOUNDED WITHIN RANGE"

Performance Test Report : Attached

The above mentioned instrument has been calibrated using standards manufacturer recommended protocols, using equipment's having traceability to National/International Standards. Devices for which there are no National / International Standards calibration standards are measured against in-house performance standard using accepted test procedures.

This report relates only to the item(s) submitted. The partial reproduction of the certificate is not permitted without approvals of M/S. Flair Biomend Engineering Pvt Ltd. The Results Reported are valid at the time of measurement.

Calibration Done Date: 11/07/2022

Calibration Due Date: 10/07/2023

Calibration Done By: SURAJ



Approved By: VIJAYA BHASKAR

Flair Biomend Engineering Private Limited

Address : H.No.2-1-1/30, Saraswati Nagar, Uppal, Hyderabad-500 039.

Mobile : +91 8008677758 | Email : info@flairbiomend.com | Web : www.flairbiomend.com

Flair Biomend Engineering Private Limited



SPEED TEST REPORT

Customer Name	ICTC LAB DIST. HQRS. Hospital, karimnagar	Equip Name	CENTRIFUGE
Serial number	1072	Model	NA
Manufacturer	BAROMETER	Department	LAB

Calibration Standards used:

Test Equipment	Manufacturer	Model	Serial No.	Cal Date	Due Date
DIGITAL TACHO METER	AMPROBE	TACH-10	63215	23.06.2022	23.06.2023

Test Results Table

S.No.	Test Gadget Set Value (RPM)	DUT Measured Value (RPM)	DEVIATION
1	1000	998	-2
2	2000	1998	-2
3	3000	2997	-3
4	4000	3998	-2

Unit Passed

Unit Failed

Remarks (If any) :

Engineers Sign:

Date: 11.07.2022

Flair Biomend Engineering Private Limited

Address: H.No:2-1-1/30, Saraswati Nagar, Uppal, Hyderabad-500 039.
Mobile:+918008677758 |Email: info@flairbiomend.com |Web:www.flairbiomend.com