



CERTIFICATE OF CALIBRATION

Issued To/Hospital Name: ICTC LAB DIST. HQRS. HOSPITAL, KARIMNAGAR.

Certificate No	FBM/C22/4325	Equipment Name	REFRIGERATOR
Serial Number	NA	Model	CLASSIC
Manufacturer	GODREJ	Department	LAB

Calibration Standards used:

S.No	Test Equipment	Manufacturer	Model No	Serial No	Cal Done Date	Cal Due Date
1	TEMPERATURE ANALYZER	HTC	IRX-64	2017084062	23.06.2022	23.06.2023

"PERFORMANCE OF THE EQUIPMENT FOUNDED WITHIN RANGE"

Performance Test Report : Attached

The above mentioned instrument has been calibrated using standards manufacturer recommended protocols, using equipment's having traceability to National/International Standards. Devices for which there are no National / International Standards calibration standards are measured against in-house performance standard using accepted test procedures.

This report relates only to the item(s) submitted. The partial reproduction of the certificate is not permitted without approvals of M/S. Flair Biomend Engineering Pvt Ltd. The Results Reported are valid at the time of measurement.

Calibration Done Date: 11/07/2022

Calibration Done By: SURAJ



Calibration Due Date: 10/07/2023

Approved By: VIJAYA BHASKAR

Flair Biomend Engineering Private Limited

Address : H.No.2-1-1/30, Saraswati Nagar, Uppal, Hyderabad-500 039.
Mobile : +91 8008677758 | Email : info@flairbiomend.com | Web : www.flairbiomend.com

Flair Biomend Engineering Private Limited



TEMPERATURE TEST REPORT

Customer Name	ICTC LAB DIST. HQRS. Hospital, karimnagar	Equip name	REFRIGERATOR
Serial number	NA	Model	CLASSIC
Manufacturer	GODREJ	Department	LAB

Calibration Standards used:

Test Equipment	Manufacturer	Model	Serial No.	Cal Date	Due Date
Temperature Analyzer	HTC	IRX-64	2017084062	23.06.2022	23.06.2023

Test Results Table

S/no	Test Gadget Set Value (°C)	DUT Measured Value (°C)	DEVIATION
1	2	2.2	0.2
2	5	5.1	0.1
3	8	8.1	0.1

Unit Passed

Unit Failed

Remarks (If any) :

Engineers Sign:

Date: 11.07.2022

Flair Biomend Engineering Private Limited

Address: H.No.2-1-1/30, Saraswati Nagar, Uppal, Hyderabad-500 039.
Mobile:+918008677758 |Email: info@flairbiomend.com |Web:www.flairbiomend.com