

23-12-2022 TO 23-12-2023

To

BEST DIAGNO LAB
NO.1B,Sai Guru kripa,1st Floor
NO.47/11,East Mada Street
Thiruvanmiyur,Chennai-41

CALIBRATION CERTIFICATE

Instrument : Electrolyte Analyzer
Model : UNILYTE 6+
Serial No : 2070
Make : SensaCore, India

This is certify that the UNILYTE 6+ has undergoes Automatic Calibration regularly and hence manual calibration is not required. Instrument is working with the reliable accuracy limits.

Thanking you,

Yours Truly,

For DiaReK HealthCare (P) Ltd.,



P. RAHUL
Sr. Service Engineer
For DiaReK Healthcare Pvt Ltd.

DiaReK Health Care Pvt. Ltd.

INTERNAL QUALITY ASSURANCE LOG REGISTER

REV. DATE: 01-12-2022

REV NO: 01

FORMAT NO: F/BDL/16

INTERNAL QUALITY ASSURANCE		BIO CHEMISTRY		DATE		19-12-2022		TIME		10.02 AM	
EQUIPEMENT NAME		LOT NO		RESULT		UNIT		MEAN VALUE		SD VALUE	
S.NO	SPECIFIC TESTS/EXAMINATION PERFORMED									QA RUN BY	QA MONITOR BY
1.	GLUCOSE	2007	104	mg/dL	111	8.75				S. AMMU	R. DE UMASHANKAR
2.	UREA	2002	43.71	mg/dL	43.9	3.35				S. AMMU	R. DE UMASHANKAR
3.	CREATININE	2011	1.32	mg/dL	1.46	0.16				S. AMMU	R. DE UMASHANKAR
4.	T.BILIRUBIN	2003	1.47	mg/dL	1.68	0.17				S. AMMU	R. DE UMASHANKAR
5.	T-PROTEIN	1007	5.5	g/dL	5.74	0.61				S. AMMU	R. DE UMASHANKAR
6.	ALBUMIN	1008	3.7	3.7g/dL	4.12	0.31				S. AMMU	R. DE UMASHANKAR
7.	CALCIUM	1003	9.22	mg/dL	8.74	0.44				S. AMMU	R. DE UMASHANKAR
8.	PHOSPHORUS	1001	4.40	mg/dL	4.15	0.34				S. AMMU	R. DE UMASHANKAR
9.	URIC ACID	1007	5.61	mg/dL	5.83	0.41				S. AMMU	R. DE UMASHANKAR
10.	CHOLESTEROL	2006	155.2	mg/dL	153	10.5				S. AMMU	R. DE UMASHANKAR
11.	TRIGLYCERIDE	2002	99.3	mg/dL	96.5	7.35				S. AMMU	R. DE UMASHANKAR
12.	HDL	2010	45.6	mg/dL	54.4	3.85				S. AMMU	R. DE UMASHANKAR
13.	SODIUM	ST-100 B	148.6	mmol/L	150	0.98				S. AMMU	R. DE UMASHANKAR
14.	POTASSIUM	ST-100 B	3.5	mmol/L	3.3	0.14				S. AMMU	R. DE UMASHANKAR
15.	CHLORIDE	ST-100 B	110	mmol/L	108	1.4				S. AMMU	R. DE UMASHANKAR
16.	ALP	1205	177.2	U/L	203	13				S. AMMU	R. DE UMASHANKAR

Mindray BS-240 Fully Auto Analyzer/Electrolyte - DIO REK


