

Mindray BS-240 Auto Chemistry Analyzer – CC Documents

Customer : **Best Diagno lab** Contact person : Mr.Umashankar.R.DE
Address : 1B "Sai Guru Kripa" 1st floor, Rear block, Telephone no : 9840624076
no:47, East mada street, Thiruvanmiyur
Chennai- 600041

Serial no : **YX-89001947** Performed on : 11/10/2022

Calibration Certificate

CALIBRATION:

The Analyzer has been calibrated with Mindray Multi Sera Calibrator for Glucose, Urea, Creatinine, Uric acid, Calcium, Phosphorous, T.Bilirubin, D.Bilirubin, AST, ALT, ALP, GGT, Total protein, Albumin, Total cholesterol, Triglycerides, HDL-C.

Calibrator Manufacturer : Mindray
Calibrator name : Multi Sera Calibrator
Lot Numer : 150121009
Expiry date : 31.12.2023


Carried out Multi sera Calibrator found values are within specified range.

REMARKS

The system meets all criteria outlined in this calibration certificate documents. All the required parameter have been checked and instrument is ready for regular use.

M. Srikanth

Performer : **Mr. Srikanth.M**
Designation : Application specialist
Date : 28/12/2024



Reviewer : **Mr.MuthuVigneshwar . K.P.**
Designation : Application Manager



INTERNAL QUALITY ASSURANCE LOG REGISTER

REV. DATE: 01-12-2022

REV NO: 01

FORMAT NO: F/BDL/16

INTERNAL QUALITY ASSURANCE		BIO CHEMISTRY		DATE		19-12-2022		TIME		10.02 AM					
EQUIPEMENT NAME		LOT NO		RESULT		UNIT		MEAN VALUE		SD VALUE		QA RUN BY		QA MONITOR BY	
S.NO	SPECIFIC TESTS/EXAMINATION PERFORMED														
1.	GLUCOSE	2007	104	mg/dL	111	8.75	R. DE UMASHANKAR	S. AMMU							
2.	UREA	2002	43.71	mg/dL	43.9	3.35	R. DE UMASHANKAR	S. AMMU							
3.	CREATININE	2011	1.32	mg/dL	1.46	0.16	R. DE UMASHANKAR	S. AMMU							
4.	T.BILIRUBIN	2003	1.47	mg/dL	1.68	0.17	R. DE UMASHANKAR	S. AMMU							
5.	T-PROTEIN	1007	5.5	g/dL	5.74	0.61	R. DE UMASHANKAR	S. AMMU							
6.	ALBUMIN	1008	3.7	3.7g/dL	4.12	0.31	R. DE UMASHANKAR	S. AMMU							
7.	CALCIUM	1003	9.22	mg/dL	8.74	0.44	R. DE UMASHANKAR	S. AMMU							
8.	PHOSPHORUS	1001	4.40	mg/dL	4.15	0.34	R. DE UMASHANKAR	S. AMMU							
9.	URIC ACID	1007	5.61	mg/dL	5.83	0.41	R. DE UMASHANKAR	S. AMMU							
10.	CHOLESTEROL	2006	155.2	mg/dL	153	10.5	R. DE UMASHANKAR	S. AMMU							
11.	TRIGLYCERIDE	2002	99.3	mg/dL	96.5	7.35	R. DE UMASHANKAR	S. AMMU							
12.	HDL	2010	45.6	mg/dL	54.4	3.85	R. DE UMASHANKAR	S. AMMU							
13.	SODIUM	ST-100 B	148.6	mmol/L	150	0.98	R. DE UMASHANKAR	S. AMMU							
14.	POTASSIUM	ST-100 B	3.5	mmol/L	3.3	0.14	R. DE UMASHANKAR	S. AMMU							
15.	CHLORIDE	ST-100 B	110	mmol/L	108	1.4	R. DE UMASHANKAR	S. AMMU							
16.	ALP	1205	177.2	U/L	203	13	R. DE UMASHANKAR	S. AMMU							

Mindray BS-240 Fully Auto Analyzer/Electrolyte - DIALYK

BEST Diagno Lab
 19, Sai Guru Kripa,
 No. 47 (O.No.14) East Mada Street,
 Thiruvannamur, Chennai - 600 041.
 Ph: 24451850 Cell: 9940151650