

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company  
No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCA8053H1ZR • DL No. MDU/4205/20B, MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Guru Micro Lab, #Near police station, Salem police station, Komarapalayam-638183	Certificate No:	AMS/LAB/289/2051
		Issuing Date:	03/09/2022
		Page No:	1 of 1

<b>Details of Device under Calibration (DUC):</b>		Id. No:	GML/LAB/VMP/001
DUC:	Micropipette	Department:	Laboratory
Make:	Dragon Lab	Calibrated On:	02/09/2022
Model:	---	Calibration Due Suggested:	01/09/2023
Sl. No:	YE213AS0168303	DUC Condition on Receipt:	OK
Type:	Variable		
Range & Least Count:	5-50µl		

<b>Traceability Details:</b>				Valid Up to
Sl.No.	Test Equipment	Make	Certificate Number	
1	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

(At the time of Calibration)

Room Temperature: 25°C±2 Room Humidity: 30-60%

### Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	5.0	04.50	04.50	-0.50
2	10.0	09.00	09.00	-1.00
3	20.0	18.00	18.00	-2.00
4	30.0	28.00	28.00	-2.00
5	50.0	50.00	50.00	-0.00

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

**Calibrated By:**

(Calibration Engineer)



**Approved by:**

(Technical Manager)

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company  
No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.  
Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Guru Micro Lab, #Near police station, Salem police station, Komarapalayam-638183	Certificate No:	AMS/LAB/289/2053
		Issuing Date:	03/09/2022
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	GML/LAB/VMP/003
Make:	Dragon Lab	Department:	Laboratory
Model:	---	Calibrated On:	02/09/2022
Sl. No:	YEA13AD0037482	Calibration Due Suggested:	01/09/2023
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	100 - 1000µl		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023


Room Temperature: 25°C±2 Room Humidity: 30-60%

(At the time of Calibration)

### Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	100.0	101.00	101.00	+1.00
2	200.0	210.00	210.00	+10.00
3	500.0	500.00	500.00	+00.00
4	1000.0	1000.00	1000.00	0.00

Allowable Tolerance: User Defined



### Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)



www.acemedical.in

### Approved by:

(Technical Manager)

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company  
No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.  
Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCAS053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Guru Micro Lab, Near police station, Salem police station, Komarapalayam-638183	Certificate No:	AMS/LAB/289/2052
		Issuing Date:	03/09/2022
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	GML/LAB/VMP/002
Make:	Dragon Lab	Department:	Laboratory
Model:	---	Calibrated On:	02/09/2022
Sl. No:	YEA18AAL0147873	Calibration Due Suggested:	01/09/2023
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	100 - 1000ul		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

### Test Results Volume ul:

SL. No.	DUC Reading (ul)	Standard Reading (ul)		Correction (ul)
		Trial 1	Trial 2	
1	100.0	101.00	101.00	+1.00
2	200.0	210.00	210.00	+10.00
3	500.0	500.00	500.00	+00.00
4	1000.0	1000.00	1000.00	0.00


Allowable Tolerance: User Defined

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)



Approved by:

(Technical Manager)