



Calibrator Calibration

Instrument: 5 Part Haematology Analyzer		Model: BC -6200		Serial No.: TW-03000988		Date of Cal. 26.07.2022
Mode: OV-WB		Calibrator Lot No. PLUS0722			Exp.Date:05.08.2022	
Parameter	WBC	RBC	HGB	MCV	PLT	WBC-D
TARGET	9.05	4.59	13.6	92.5	255	9.05
1.	8.60	4.67	14.5	92.2	250	8.45
2.	8.82	4.53	14.5	92.1	241	8.67
3.	8.85	4.53	14.5	92.0	235	8.48
4.	8.49	4.58	14.4	91.7	247	8.46
5.	8.87	4.64	14.6	91.5	247	8.54
6.	8.79	4.61	14.5	91.8	254	8.49
Mean	8.74	4.59	14.5	91.9	246	8.52
CV(%)	1.773	1.281	0.38	0.31	2.8	0.973
New Cal. Factor (%)	103.59	99.91	93.87	100.64	103.80	
Old Cal. Factors (%)	103.92	95.66	93.74	99.62	102.62	
Old TC1 (%)						97.90
New TC1 (%)						102.59

For Rapid Diagnostic Pvt. Ltd.


K. SRIKANTH
 Service Engineer


N. Mahender Reddy
 Zonal Service Manager



Calibration Certificate

Customer Name	Telangana Diagnostic Hub KHAMMAM
Address	Govt. Hospital, KHAMMAM-507001
Contact Person	BINDHU
Contact No.	9985568808
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-03000988
Instrument Status	Under Warranty
Calibration Done On	26.07.2022
Next Calibration Due On	25.07.2023

For Rapid Diagnostic Pvt. Ltd.


K. SRIKANTH
Service Engineer


N. Mahender Reddy
Zonal Service Manager