

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Albumin, g/dL**

**Comparative Statistics**


	N	Mean	SD	CV%	Excluded
All Methods	280	5.77	0.60	10.35	9
Bromcresol Green (BCG) - (Your Method)	276	5.78	0.59	10.25	8
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

<b>Your Result</b>	5.67	<b>SDI / z-Score</b>	0.41
<b>Assigned Value</b>	5.48	<b>Average SDI</b>	0.05
<b>SDPA</b>	0.470	<b>Uncertainty of AV</b>	0.09
<b>Standard Unit Result Val</b>	5.67	<b>%DEV</b>	-1.84


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, Albumin**

	N	Mean	SD	CV%	Excluded
All Analyzer	280	5.77	0.60	10.35	9
<b>Your Analyzer : Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
Fuji Dri-Chem NX500/i	35	6.01	0.41	6.87	1
Roche cobas c 311	28	6.04	0.20	3.28	--
Erba EM 200	23	5.15	0.33	6.48	1
Roche cobas c 111	20	6.03	0.47	7.73	--
Biosystems A15	16	5.90	0.65	11.05	2
Fuji Dri-Chem 4000i	15	5.91	0.30	5.08	1
Fuji Dri-Chem NX600	15	5.96	0.20	3.32	1
Roche cobas INTEGRA	13	6.31	0.42	6.61	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, ALP, U/L**

**Comparative Statistics**


	N	Mean	SD	CV%	Excluded
All Methods	283	54.88	13.52	24.63	3
PNPP, AMP Buffer - (Your Method)	155	54.76	11.10	20.26	2
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Method wise Peer Grouping**

Your Result	50.00	SDI / z-Score	-0.25
Assigned Value	53.21	Average SDI	-0.34
SDPA	12.700	Uncertainty of AV	0.62
Standard Unit Result Val	50.00	%DEV	-8.69

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, ALP		N	Mean	SD	CV%	Excluded
	All Analyzer	283	54.88	13.52	24.63	3
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	40	57.64	9.19	15.94	1
	Roche cobas c 311	28	51.42	8.16	15.87	--
	Roche cobas c 111	21	51.47	6.42	12.48	--
	Erba EM 200	20	49.58	7.91	15.96	1
	Biosystems A15	15	54.66	14.72	26.94	1
	Fuji Dri-Chem 4000i	15	53.73	8.60	16.00	--
	Fuji Dri-Chem NX600	15	51.00	5.43	10.64	--
	Roche cobas INTEGRA	13	57.08	5.57	9.76	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, ALT, U/L**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	287	11.18	2.75	24.63	11
UV without P5P - (Your Method)	99	9.71	1.92	19.80	1
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	10.40	SDI / z-Score	0.02
Assigned Value	10.36	Average SDI	0.40
SDPA	1.549	Uncertainty of AV	0.32
Standard Unit Result Val	10.40	%DEV	7.16


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, ALT		N	Mean	SD	CV%	Excluded
	All Analyzer	287	11.18	2.75	24.63	11
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	37	14.00	0.91	6.52	--
	Roche cobas c 311	28	10.44	1.37	13.13	--
	Erba EM 200	23	9.77	2.76	28.20	1
	Roche cobas c 111	22	8.78	0.90	10.26	--
	Biosystems A15	16	13.96	3.29	23.57	1
	Fuji Dri-Chem NX600	16	13.19	1.17	8.85	--
	Fuji Dri-Chem 4000i	15	13.67	1.45	10.59	--
	Roche cobas INTEGRA	13	7.90	1.06	13.40	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, AST, U/L**

**Comparative Statistics**



	N	Mean	SD	CV%	Excluded
All Methods	282	19.91	3.75	18.86	5
UV without P5P - (Your Method)	105	17.93	2.54	14.16	1
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	21.50	SDI / z-Score	0.69
Assigned Value	19.79	Average SDI	0.14
SDPA	2.475	Uncertainty of AV	0.58
Standard Unit Result Val	21.50	%DEV	19.91

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, AST**

	N	Mean	SD	CV%	Excluded
All Analyzer	282	19.91	3.75	18.86	5
<b>Your Analyzer : Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
Fuji Dri-Chem NX500/i	37	21.70	1.45	6.68	--
Roche cobas c 311	27	17.94	1.68	9.36	--
Erba EM 200	22	17.31	3.15	18.23	--
Roche cobas c 111	22	17.31	1.30	7.53	--
Biosystems A15	16	22.63	4.48	19.81	--
Fuji Dri-Chem 4000i	15	21.40	1.68	7.86	--
Fuji Dri-Chem NX600	15	21.87	1.25	5.70	--
Roche cobas INTEGRA	13	16.13	1.32	8.15	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error





**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Bilirubin Direct, mg/dl**

**Comparative Statistics**


	N	Mean	SD	CV%	Excluded
All Methods	273	0.48	0.48	101.90	20
Diazotization - (Your Method)	202	0.49	0.54	110.06	11
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	0.81	SDI / z-Score	-0.33
Assigned Value	0.86	Average SDI	0.78
SDPA	0.153	Uncertainty of AV	0.06
Standard Unit Result Val	0.81	%DEV	66.32

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Bilirubin Direct		N	Mean	SD	CV%	Excluded
	All Analyzer	273	0.48	0.48	101.47	20
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	39	0.10	0.03	29.29	1
	Roche cobas c 311	26	0.55	0.11	20.62	--
	Roche cobas c 111	22	0.43	0.08	17.55	--
	Erba EM 200	20	0.38	0.11	29.79	--
	Biosystems A15	16	1.89	1.00	52.64	2
	Fuji Dri-Chem 4000i	15	0.10	0.00	0.00	--
	Fuji Dri-Chem NX600	15	0.10	0.00	0.00	1
	Roche cobas INTEGRA	13	0.48	0.04	7.79	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Bilirubin Total, mg/dl**

**Comparative Statistics**



	N	Mean	SD	CV%	Excluded
All Methods	279	0.80	0.68	84.86	9
DPD - (Your Method)	10	0.61	0.23	38.17	--
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	0.90	SDI / z-Score	2.89
Assigned Value	0.47	Average SDI	2.89
SDPA	0.149	Uncertainty of AV	0.03
Standard Unit Result Val	0.90	%DEV	46.82

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Bilirubin Total		N	Mean	SD	CV%	Excluded
	All Analyzer	279	0.81	0.68	84.22	9
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	37	0.28	0.10	37.41	1
	Roche cobas c 311	27	0.95	0.16	16.95	--
	Roche cobas c 111	22	0.97	0.18	18.14	--
	Erba EM 200	21	0.39	0.28	72.38	2
	Biosystems A15	17	2.51	0.82	32.79	--
	Fuji Dri-Chem 4000i	15	0.21	0.06	28.50	--
	Fuji Dri-Chem NX600	15	0.21	0.06	30.05	--
	Roche cobas INTEGRA	13	1.08	0.07	6.79	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Calcium, mg/dl**

**Comparative Statistics**


	N	Mean	SD	CV%	Excluded
All Methods	230	8.04	0.73	9.12	2
Arsenazo III - (Your Method)	82	8.29	0.82	9.94	1
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	8.35	SDI / z-Score	0.05
Assigned Value	8.31	Average SDI	-0.16
SDPA	0.754	Uncertainty of AV	0.16
Standard Unit Result Val	8.35	%DEV	0.72

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Calcium		N	Mean	SD	CV%	Excluded
	All Analyzer	230	8.04	0.73	9.12	2
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	36	7.86	0.56	7.18	1
	Roche cobas c 311	27	7.92	0.27	3.38	--
	Roche cobas c 111	20	7.74	0.34	4.39	--
	Erba EM 200	15	8.28	0.78	9.40	--
	Fuji Dri-Chem 4000i	15	7.89	0.79	10.02	--
	Fuji Dri-Chem NX600	15	7.71	0.41	5.27	--
	Roche cobas INTEGRA	13	7.96	0.27	3.37	--
	Roche cobas 6000	10	7.84	0.20	2.59	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Cholesterol, mg/dl**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	280	148.41	11.24	7.57	11
Cholesterol oxidase, esterase, peroxidase - (Your Method)	177	147.82	10.20	6.90	5
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	152.60	SDI / z-Score	-0.05
Assigned Value	153.29	Average SDI	-0.01
SDPA	13.600	Uncertainty of AV	2.14
Standard Unit Result Val	152.60	%DEV	3.24


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, Cholesterol**

	N	Mean	SD	CV%	Excluded
All Analyzer	280	148.41	11.24	7.57	11
<b>Your Analyzer : Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
Fuji Dri-Chem NX500/i	34	145.88	12.07	8.27	1
Roche cobas c 311	28	149.08	4.79	3.21	--
Roche cobas c 111	23	151.91	10.42	6.86	--
Erba EM 200	22	148.71	14.09	9.48	--
Biosystems A15	16	154.30	13.39	8.68	3
Fuji Dri-Chem 4000i	15	144.80	9.31	6.43	1
Fuji Dri-Chem NX600	14	148.36	6.61	4.45	--
Roche cobas INTEGRA	13	149.53	3.89	2.60	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error





**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Creatinine, mg/dl**


**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	284	7.01	0.67	9.58	9
Alkaline picrate-kinetic - (Your Method)	48	6.95	0.60	8.64	5
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

<b>Your Result</b>	7.42	<b>SDI / z-Score</b>	0.62
<b>Assigned Value</b>	6.92	<b>Average SDI</b>	0.39
<b>SDPA</b>	0.809	<b>Uncertainty of AV</b>	0.22
<b>Standard Unit Result Val</b>	7.42	<b>%DEV</b>	6.79

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Creatinine		N	Mean	SD	CV%	Excluded
	All Analyzer	284	7.01	0.67	9.58	9
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	35	7.14	0.44	6.21	--
	Roche cobas c 311	28	6.97	0.38	5.51	--
	Erba EM 200	23	7.19	0.52	7.19	1
	Roche cobas c 111	23	6.79	0.39	5.77	--
	Biosystems A15	16	6.77	0.85	12.59	--
	Fuji Dri-Chem 4000i	16	7.47	0.55	7.37	1
	Fuji Dri-Chem NX600	14	6.92	0.37	5.39	--
	Roche cobas INTEGRA	13	6.83	0.33	4.83	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, GGT, U/L**

**Comparative Statistics**



	N	Mean	SD	CV%	Excluded
All Methods	203	28.44	2.81	9.87	1
IFCC - (Your Method)	18	29.64	2.13	7.17	--
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	29.00	SDI / z-Score	0.32
Assigned Value	28.20	Average SDI	0.59
SDPA	2.500	Uncertainty of AV	0.24
Standard Unit Result Val	29.00	%DEV	-2.17

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, GGT		N	Mean	SD	CV%	Excluded
	All Analyzer	203	28.44	2.81	9.87	1
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	32	29.59	2.16	7.28	1
	Roche cobas c 311	28	28.29	1.46	5.17	--
	Roche cobas c 111	21	26.87	2.31	8.58	--
	Fuji Dri-Chem 4000i	13	29.31	1.65	5.64	--
	Roche cobas INTEGRA	13	27.69	2.00	7.22	--
	Fuji Dri-Chem NX600	13	26.92	1.04	3.86	--
	Erba EM 200	9	28.30	3.48	12.29	--
	Roche cobas 6000	8	28.59	1.68	5.87	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Glucose, mg/dl**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	282	245.40	22.52	9.18	9
Glucose oxidase, hydrogen peroxide (Trinder) - (Your Metho	127	240.75	27.22	11.31	6
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

<b>Your Result</b>	251.00	<b>SDI / z-Score</b>	-0.24
<b>Assigned Value</b>	254.40	<b>Average SDI</b>	-0.12
<b>SDPA</b>	14.457	<b>Uncertainty of AV</b>	2.61
<b>Standard Unit Result Val</b>	251.00	<b>%DEV</b>	4.26

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Glucose		N	Mean	SD	CV%	Excluded
	All Analyzer	282	245.40	22.52	9.18	9
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	37	217.25	6.60	3.04	1
	Roche cobas c 311	28	250.43	8.67	3.46	--
	Erba EM 200	23	272.66	18.41	6.75	1
	Roche cobas c 111	20	255.93	14.56	5.69	--
	Biosystems A15	16	257.38	17.81	6.92	--
	Fuji Dri-Chem 4000i	16	212.87	12.27	5.76	1
	Fuji Dri-Chem NX600	14	214.86	8.28	3.86	--
	Roche cobas INTEGRA	13	253.66	9.79	3.86	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, HDL, mg/dl**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	261	31.83	5.23	16.43	9
Direct measure, immunoinhibition - (Your Method)	3	30.97	0.90	2.89	--
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

<b>Your Result</b>	32.00	<b>SDI / z-Score</b>	0.07
<b>Assigned Value</b>	31.66	<b>Average SDI</b>	-0.26
<b>SDPA</b>	4.667	<b>Uncertainty of AV</b>	0.71
<b>Standard Unit Result Val</b>	32.00	<b>%DEV</b>	3.34

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, HDL		N	Mean	SD	CV%	Excluded
	All Analyzer	261	31.83	5.23	16.43	9
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	32	31.00	2.93	9.44	--
	Roche cobas c 311	28	29.16	1.78	6.12	--
	Erba EM 200	22	30.60	4.32	14.13	2
	Roche cobas c 111	22	29.28	1.35	4.61	--
	Fuji Dri-Chem 4000i	14	31.57	3.30	10.45	--
	Biosystems A15	13	43.25	7.74	17.88	1
	Fuji Dri-Chem NX600	13	30.54	1.33	4.36	--
	Roche cobas INTEGRA	12	29.09	1.45	4.98	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error





**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Phosphorous Inorganic, mg/dl**

**Comparative Statistics**


	N	Mean	SD	CV%	Excluded
All Methods	186	3.92	0.35	8.98	1
Phosphomolybdate method - (Your Method)	95	3.89	0.36	9.25	1
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	3.55	SDI / z-Score	-0.52
Assigned Value	3.71	Average SDI	-0.47
SDPA	0.318	Uncertainty of AV	0.09
Standard Unit Result Val	3.55	%DEV	-8.83

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Phosphorous Inorganic		N	Mean	SD	CV%	Excluded
	All Analyzer	186	3.92	0.35	8.98	1
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	36	3.88	0.16	4.04	--
	Roche cobas c 311	24	3.77	0.09	2.26	--
	Fuji Dri-Chem 4000i	14	3.80	0.14	3.71	--
	Fuji Dri-Chem NX600	13	3.87	0.11	2.87	--
	Roche cobas c 111	12	3.82	0.18	4.76	--
	Roche cobas INTEGRA	11	3.85	0.15	3.90	--
	Roche cobas 6000	10	3.74	0.14	3.74	--
	Erba EM 200	8	4.15	0.46	11.17	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Potassium, mmol/L**

**Comparative Statistics**


	N	Mean	SD	CV%	Excluded
All Methods	227	4.14	0.23	5.65	2
ISE indirect - (Your Method)	57	4.11	0.19	4.72	1
Allcare AC-9801 - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Method wise Peer Grouping**

Your Result	3.58	SDI / z-Score	-2.72
Assigned Value	4.13	Average SDI	-2.44
SDPA	0.200	Uncertainty of AV	0.02
Standard Unit Result Val	3.58	%DEV	-12.90

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Potassium		N	Mean	SD	CV%	Excluded
	All Analyzer	227	4.14	0.23	5.65	2
<b>Your Analyzer :</b>	<b>Allcare AC-9801</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	37	4.13	0.26	6.28	1
	Roche cobas c 311	21	4.17	0.12	2.78	--
	Fuji Dri-Chem 4000i	19	4.19	0.29	6.94	--
	Roche AVL 9180	19	4.25	0.09	2.12	--
	Fuji Dri-Chem NX600	13	4.14	0.10	2.32	--
	EasyLyte	10	4.11	0.20	4.77	--
	BeneSphera Electrolyte Analyzer	9	3.92	0.16	4.11	--
	Roche cobas 6000	8	4.11	0.10	2.43	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Sodium, mmol/L**

**Comparative Statistics**



	N	Mean	SD	CV%	Excluded
All Methods	224	158.12	5.92	3.74	6
ISE indirect - (Your Method)	58	157.95	4.63	2.93	2
Allcare AC-9801 - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Method wise Peer Grouping**

Your Result	155.40	SDI / z-Score	-0.57
Assigned Value	157.86	Average SDI	-0.78
SDPA	4.332	Uncertainty of AV	0.72
Standard Unit Result Val	155.40	%DEV	-1.61

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Sodium		N	Mean	SD	CV%	Excluded
	All Analyzer	224	158.12	5.92	3.74	6
<b>Your Analyzer :</b>	<b>Allcare AC-9801</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	37	158.79	6.46	4.07	3
	Roche cobas c 311	21	158.64	3.99	2.52	--
	Roche AVL 9180	19	161.10	3.90	2.42	1
	Fuji Dri-Chem 4000i	18	158.33	6.64	4.20	--
	Fuji Dri-Chem NX600	12	159.25	2.70	1.70	--
	EasyLyte	10	157.43	6.10	3.87	--
	BeneSphera Electrolyte Analyzer	9	150.32	4.12	2.74	--
	Roche cobas 6000	8	156.88	3.00	1.91	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Total Protein, g/dL**

**Comparative Statistics**



	N	Mean	SD	CV%	Excluded
All Methods	279	8.78	0.58	6.63	7
Biuret - (Your Method)	271	8.79	0.59	6.69	6
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	8.56	SDI / z-Score	-0.07
Assigned Value	8.61	Average SDI	0.11
SDPA	0.620	Uncertainty of AV	0.09
Standard Unit Result Val	8.56	%DEV	-2.57

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Total Protein		N	Mean	SD	CV%	Excluded
	All Analyzer	279	8.78	0.58	6.63	7
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	38	8.88	0.51	5.79	1
	Roche cobas c 311	28	8.73	0.18	2.03	--
	Erba EM 200	23	8.93	0.68	7.59	--
	Roche cobas c 111	19	8.58	0.27	3.14	--
	Biosystems A15	16	9.07	0.70	7.71	1
	Fuji Dri-Chem 4000i	15	9.03	0.56	6.17	1
	Fuji Dri-Chem NX600	14	8.64	0.18	2.12	--
	Roche cobas INTEGRA	13	8.45	0.23	2.75	--

Remarks: Interpretation of your reports

- If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error





**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Triglycerides, mg/dl**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	272	128.85	14.51	11.26	9
Enzymatic, end point - (Your Method)	61	135.19	15.02	11.11	1
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	130.10	SDI / z-Score	0.01
Assigned Value	129.97	Average SDI	-0.07
SDPA	9.500	Uncertainty of AV	0.70
Standard Unit Result Val	130.10	%DEV	-3.76


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



PC-1022

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, Triglycerides**

	N	Mean	SD	CV%	Excluded
All Analyzer	272	128.85	14.51	11.26	9
<b>Your Analyzer : Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
Fuji Dri-Chem NX500/i	32	137.77	16.24	11.79	--
Roche cobas c 311	28	130.28	3.31	2.54	--
Roche cobas c 111	21	135.99	9.72	7.15	2
Erba EM 200	20	110.46	8.80	7.97	--
Biosystems A15	16	121.29	12.81	10.56	1
Fuji Dri-Chem 4000i	14	135.07	10.91	8.08	--
Roche cobas INTEGRA	13	131.32	5.43	4.13	--
Fuji Dri-Chem NX600	13	132.92	5.88	4.42	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Urea, mg/dl**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	283	22.55	2.15	9.52	18
Urease, UV - (Your Method)	48	22.56	2.35	10.42	6
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

<b>Your Result</b>	21.20	<b>SDI / z-Score</b>	-0.74
<b>Assigned Value</b>	22.47	<b>Average SDI</b>	-0.76
<b>SDPA</b>	1.726	<b>Uncertainty of AV</b>	0.47
<b>Standard Unit Result Val</b>	21.20	<b>%DEV</b>	-6.02

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Urea		N	Mean	SD	CV%	Excluded
	All Analyzer	283	22.55	2.15	9.52	18
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	35	22.97	1.61	7.00	--
	Roche cobas c 311	28	22.99	1.26	5.48	1
	Erba EM 200	23	23.65	2.83	11.95	2
	Roche cobas c 111	22	21.25	1.39	6.54	1
	Biosystems A15	16	21.61	3.21	14.83	1
	Fuji Dri-Chem 4000i	15	22.98	2.39	10.41	--
	Fuji Dri-Chem NX600	14	22.98	0.69	2.99	1
	Roche cobas INTEGRA	13	21.66	1.48	6.83	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Uric Acid, mg/dl**

**Comparative Statistics**


	N	Mean	SD	CV%	Excluded
All Methods	275	4.40	0.71	16.13	10
Uricase, colorimetric - (Your Method)	155	4.32	0.60	13.83	4
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Equipment wise Peer Grouping**

<b>Your Result</b>	4.10	<b>SDI / z-Score</b>	0.36
<b>Assigned Value</b>	3.97	<b>Average SDI</b>	0.38
<b>SDPA</b>	0.360	<b>Uncertainty of AV</b>	0.02
<b>Standard Unit Result Val</b>	4.10	<b>%DEV</b>	-5.00

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, Uric Acid**

	N	Mean	SD	CV%	Excluded
All Analyzer	275	4.40	0.71	16.13	10
<b>Your Analyzer : Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
Fuji Dri-Chem NX500/i	33	4.53	0.26	5.64	2
Roche cobas c 311	26	3.95	0.23	5.90	--
Erba EM 200	21	4.65	0.41	8.84	2
Roche cobas c 111	21	3.97	0.19	4.66	1
Fuji Dri-Chem 4000i	18	4.44	0.26	5.81	2
Biosystems A15	16	5.03	0.57	11.22	--
Fuji Dri-Chem NX600	14	4.47	0.15	3.33	--
Roche cobas INTEGRA	13	4.10	0.32	7.69	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416,  
Commercial Building - 1A,  
Kohinoor Mall, Kirol Road, Kurla (W),  
Mumbai - 400070



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Albumin, g/dL**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	285	4.17	0.50	11.99	7
Bromcresol Green (BCG) - (Your Method)	281	4.17	0.50	11.92	7
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

<b>Your Result</b>	3.75	<b>SDI / z-Score</b>	-0.31
<b>Assigned Value</b>	3.84	<b>Average SDI</b>	0.05
<b>SDPA</b>	0.278	<b>Uncertainty of AV</b>	0.05
<b>Standard Unit Result Val</b>	3.75	<b>%DEV</b>	-10.09

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, Albumin**

	N	Mean	SD	CV%	Excluded
All Analyzer	285	4.17	0.50	11.99	7
<b>Your Analyzer : Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
Fuji Dri-Chem NX500/i	38	4.88	0.40	8.21	2
Roche cobas c 311	29	3.94	0.15	3.88	--
Erba EM 200	23	3.94	0.18	4.67	--
Roche cobas c 111	21	4.09	0.18	4.45	--
Biosystems A15	16	3.85	0.38	9.75	2
Fuji Dri-Chem NX600	16	4.56	0.20	4.36	1
Fuji Dri-Chem 4000i	15	4.91	0.40	8.13	--
Roche cobas INTEGRA	13	3.99	0.18	4.48	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error





**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, ALP, U/L**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	283	56.84	13.54	23.82	2
PNPP, AMP Buffer - (Your Method)	154	56.65	9.85	17.39	1
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--


**Evaluation Based On - Method wise Peer Grouping**

Your Result	50.00	SDI / z-Score	-0.42
Assigned Value	55.29	Average SDI	-0.34
SDPA	12.700	Uncertainty of AV	0.58
Standard Unit Result Val	50.00	%DEV	-11.74


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, ALP		N	Mean	SD	CV%	Excluded
	All Analyzer	283	56.84	13.54	23.82	2
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	40	60.21	8.72	14.48	1
	Roche cobas c 311	29	54.76	5.08	9.27	--
	Roche cobas c 111	21	53.71	6.60	12.28	--
	Erba EM 200	20	57.16	9.40	16.45	1
	Biosystems A15	15	53.22	14.55	27.35	--
	Fuji Dri-Chem 4000i	15	55.07	5.61	10.19	--
	Fuji Dri-Chem NX600	15	52.13	5.04	9.67	--
	Roche cobas INTEGRA	13	56.90	4.65	8.17	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, ALT, U/L**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	286	8.74	2.36	26.97	18
UV without P5P - (Your Method)	97	7.54	1.60	21.20	3
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

<b>Your Result</b>	9.70	<b>SDI / z-Score</b>	0.78
<b>Assigned Value</b>	8.44	<b>Average SDI</b>	0.40
<b>SDPA</b>	1.626	<b>Uncertainty of AV</b>	0.34
<b>Standard Unit Result Val</b>	9.70	<b>%DEV</b>	28.60

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, ALT		N	Mean	SD	CV%	Excluded
	All Analyzer	286	8.74	2.36	26.97	18
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	37	10.95	1.25	11.38	--
	Roche cobas c 311	29	7.54	1.36	18.08	--
	Erba EM 200	22	7.56	1.54	20.33	4
	Roche cobas c 111	22	6.99	1.13	16.23	--
	Biosystems A15	16	10.38	2.61	25.16	2
	Fuji Dri-Chem NX600	16	10.13	1.03	10.12	--
	Fuji Dri-Chem 4000i	15	10.60	0.99	9.30	--
	Roche cobas INTEGRA	13	5.80	0.75	12.93	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, AST, U/L**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	282	16.76	3.89	23.21	6
UV without P5P - (Your Method)	104	15.13	2.73	18.06	--
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	14.90	SDI / z-Score	-0.41
Assigned Value	15.87	Average SDI	0.14
SDPA	2.394	Uncertainty of AV	0.56
Standard Unit Result Val	14.90	%DEV	-1.49


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, AST		N	Mean	SD	CV%	Excluded
	All Analyzer	282	16.76	3.89	23.21	6
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	37	18.89	0.99	5.26	--
	Roche cobas c 311	28	14.12	1.74	12.29	--
	Erba EM 200	22	15.02	3.16	21.06	--
	Roche cobas c 111	22	14.72	1.30	8.82	--
	Biosystems A15	16	18.00	5.14	28.54	--
	Fuji Dri-Chem 4000i	15	18.40	1.35	7.35	--
	Fuji Dri-Chem NX600	15	18.67	1.18	6.30	--
	Roche cobas INTEGRA	13	12.72	0.69	5.44	--

Remarks: Interpretation of your reports

- If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Bilirubin Direct, mg/dl**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	273	0.21	0.25	117.23	27
Diazotization - (Your Method)	201	0.24	0.27	112.45	17
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	0.10	SDI / z-Score	1.90
Assigned Value	0.06	Average SDI	0.78
SDPA	0.021	Uncertainty of AV	0.01
Standard Unit Result Val	0.10	%DEV	-58.51


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Bilirubin Direct		N	Mean	SD	CV%	Excluded
	All Analyzer	273	0.22	0.25	113.95	27
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	39	0.10	0.03	30.00	2
	Roche cobas c 311	27	0.38	0.08	20.47	--
	Roche cobas c 111	22	0.21	0.08	38.46	--
	Erba EM 200	19	0.08	0.05	68.42	3
	Biosystems A15	16	1.15	0.56	48.35	7
	Fuji Dri-Chem 4000i	15	0.10	0.00	0.00	--
	Fuji Dri-Chem NX600	15	0.10	0.00	0.00	--
	Roche cobas INTEGRA	13	0.20	0.03	14.36	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error





**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Calcium, mg/dl**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	230	8.29	0.59	7.07	8
Arsenazo III - (Your Method)	82	8.54	0.70	8.15	3
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	8.20	SDI / z-Score	-0.36
Assigned Value	8.40	Average SDI	-0.16
SDPA	0.555	Uncertainty of AV	0.12
Standard Unit Result Val	8.20	%DEV	-3.96


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



PC-1022

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Calcium		N	Mean	SD	CV%	Excluded
	All Analyzer	230	8.29	0.59	7.07	8
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	36	7.90	0.50	6.35	1
	Roche cobas c 311	28	8.26	0.24	2.88	--
	Roche cobas c 111	20	8.10	0.40	4.92	1
	Erba EM 200	15	8.72	0.82	9.35	--
	Fuji Dri-Chem 4000i	15	8.05	0.44	5.40	2
	Fuji Dri-Chem NX600	15	7.91	0.40	5.00	--
	Roche cobas INTEGRA	13	8.24	0.13	1.52	--
	Beckman Coulter AU480	9	8.43	0.46	5.48	--

Remarks: Interpretation of your reports

- If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Cholesterol, mg/dl**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	280	153.44	12.63	8.23	6
Cholesterol oxidase, esterase, peroxidase - (Your Method)	176	152.80	11.70	7.66	3
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	160.10	SDI / z-Score	0.04
Assigned Value	159.57	Average SDI	-0.01
SDPA	12.458	Uncertainty of AV	1.96
Standard Unit Result Val	160.10	%DEV	4.78


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



PC-1022

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, Cholesterol**

	N	Mean	SD	CV%	Excluded
All Analyzer	280	153.44	12.63	8.23	6
<b>Your Analyzer : Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
Fuji Dri-Chem NX500/i	34	142.84	10.75	7.52	--
Roche cobas c 311	29	156.22	4.52	2.89	--
Roche cobas c 111	23	156.85	10.87	6.93	--
Erba EM 200	22	154.37	12.13	7.86	--
Biosystems A15	16	157.81	17.25	10.93	1
Fuji Dri-Chem 4000i	15	147.07	12.39	8.42	1
Fuji Dri-Chem NX600	14	148.07	10.04	6.78	--
Roche cobas INTEGRA	13	157.30	4.78	3.04	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Creatinine, mg/dl**


**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	284	0.78	0.12	15.69	8
Alkaline picrate-kinetic - (Your Method)	48	0.79	0.15	18.73	3
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

<b>Your Result</b>	0.78	<b>SDI / z-Score</b>	0.15
<b>Assigned Value</b>	0.76	<b>Average SDI</b>	0.39
<b>SDPA</b>	0.121	<b>Uncertainty of AV</b>	0.03
<b>Standard Unit Result Val</b>	0.78	<b>%DEV</b>	-1.27

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



PC-1022

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Creatinine		N	Mean	SD	CV%	Excluded
	All Analyzer	284	0.78	0.12	15.69	8
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	35	0.72	0.08	10.40	--
	Roche cobas c 311	29	0.77	0.06	8.04	--
	Erba EM 200	23	0.82	0.14	16.89	--
	Roche cobas c 111	23	0.75	0.09	11.45	--
	Biosystems A15	16	0.78	0.18	23.10	2
	Fuji Dri-Chem 4000i	16	0.75	0.07	9.52	--
	Fuji Dri-Chem NX600	14	0.71	0.04	5.88	--
	Roche cobas INTEGRA	13	0.71	0.05	7.16	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, GGT, U/L**

**Comparative Statistics**



	N	Mean	SD	CV%	Excluded
All Methods	203	28.08	2.75	9.81	--
IFCC - (Your Method)	18	28.96	2.18	7.52	--
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	30.00	SDI / z-Score	0.85
Assigned Value	28.13	Average SDI	0.59
SDPA	2.200	Uncertainty of AV	0.26
Standard Unit Result Val	30.00	%DEV	3.61

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, GGT		N	Mean	SD	CV%	Excluded
	All Analyzer	203	28.08	2.75	9.81	--
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	32	29.19	2.29	7.84	--
	Roche cobas c 311	29	28.19	1.56	5.53	--
	Roche cobas c 111	21	26.62	2.12	7.95	--
	Fuji Dri-Chem 4000i	13	28.92	2.06	7.12	--
	Roche cobas INTEGRA	13	27.45	2.10	7.65	--
	Fuji Dri-Chem NX600	13	27.46	1.56	5.68	--
	Erba EM 200	10	26.38	3.64	13.78	--
	Beckman Coulter AU480	7	29.07	1.24	4.26	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error





**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Glucose, mg/dl**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	282	71.92	8.59	11.94	3
Glucose oxidase, hydrogen peroxide (Trinder) - (Your Metho	127	71.34	10.16	14.24	1
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	74.50	SDI / z-Score	0.00
Assigned Value	74.48	Average SDI	-0.12
SDPA	6.470	Uncertainty of AV	1.19
Standard Unit Result Val	74.50	%DEV	4.43


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Glucose		N	Mean	SD	CV%	Excluded
	All Analyzer	282	71.92	8.59	11.94	3
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	37	62.70	3.32	5.30	--
	Roche cobas c 311	29	71.42	2.26	3.17	--
	Erba EM 200	23	84.28	6.20	7.35	--
	Roche cobas c 111	20	74.31	4.87	6.55	--
	Biosystems A15	16	72.61	7.49	10.32	--
	Fuji Dri-Chem 4000i	16	61.69	4.60	7.46	--
	Fuji Dri-Chem NX600	14	61.00	2.80	4.59	1
	Roche cobas INTEGRA	13	72.39	2.19	3.02	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, HDL, mg/dl**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	261	32.81	5.45	16.61	5
Direct measure, immunoinhibition - (Your Method)	3	33.33	4.82	14.46	--
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

<b>Your Result</b>	30.50	<b>SDI / z-Score</b>	-0.59
<b>Assigned Value</b>	33.53	<b>Average SDI</b>	-0.26
<b>SDPA</b>	5.108	<b>Uncertainty of AV</b>	0.77
<b>Standard Unit Result Val</b>	30.50	<b>%DEV</b>	-8.50

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, HDL		N	Mean	SD	CV%	Excluded
	All Analyzer	261	32.81	5.45	16.61	5
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	32	33.71	2.19	6.51	--
	Roche cobas c 311	29	28.96	1.37	4.72	--
	Roche cobas c 111	22	29.36	1.59	5.42	--
	Erba EM 200	21	33.56	5.08	15.14	--
	Fuji Dri-Chem 4000i	14	32.92	1.44	4.38	1
	Biosystems A15	13	40.39	9.33	23.10	--
	Fuji Dri-Chem NX600	13	32.50	2.03	6.25	--
	Roche cobas INTEGRA	12	28.51	1.32	4.61	1

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Phosphorous Inorganic, mg/dl**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	185	4.06	0.38	9.42	3
Phosphomolybdate method - (Your Method)	93	4.09	0.37	9.11	3
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	3.82	SDI / z-Score	-0.42
Assigned Value	3.94	Average SDI	-0.47
SDPA	0.277	Uncertainty of AV	0.08
Standard Unit Result Val	3.82	%DEV	-6.69


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Phosphorous Inorganic		N	Mean	SD	CV%	Excluded
	All Analyzer	185	4.06	0.38	9.42	3
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	35	3.92	0.20	5.02	--
	Roche cobas c 311	25	3.93	0.09	2.27	--
	Fuji Dri-Chem 4000i	14	3.84	0.20	5.28	--
	Fuji Dri-Chem NX600	13	3.92	0.11	2.91	--
	Roche cobas c 111	12	4.09	0.21	5.17	--
	Roche cobas INTEGRA	11	3.98	0.14	3.54	--
	Roche cobas 6000	9	4.00	0.20	5.05	--
	Erba EM 200	8	4.49	0.42	9.27	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Potassium, mmol/L**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	228	4.13	0.24	5.78	4
ISE indirect - (Your Method)	57	4.12	0.19	4.51	1
Allcare AC-9801 - (Your Analyzer)	1	--	--	--	--


**Evaluation Based On - Method wise Peer Grouping**

Your Result	3.62	SDI / z-Score	-2.15
Assigned Value	4.14	Average SDI	-2.44
SDPA	0.240	Uncertainty of AV	0.02
Standard Unit Result Val	3.62	%DEV	-12.18


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Potassium		N	Mean	SD	CV%	Excluded
	All Analyzer	228	4.13	0.24	5.78	4
<b>Your Analyzer :</b>	<b>Allcare AC-9801</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	37	4.12	0.19	4.62	1
	Roche cobas c 311	22	4.15	0.17	4.12	--
	Fuji Dri-Chem 4000i	19	4.11	0.23	5.47	1
	Roche AVL 9180	19	4.23	0.14	3.24	--
	Fuji Dri-Chem NX600	14	4.14	0.09	2.27	--
	EasyLyte	10	4.04	0.15	3.59	--
	BeneSphera Electrolyte Analyzer	9	3.91	0.15	3.94	--
	Roche cobas 6000	7	4.22	0.21	4.85	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error





**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Sodium, mmol/L**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	227	149.43	5.03	3.36	4
ISE indirect - (Your Method)	58	150.25	4.96	3.30	1
Allcare AC-9801 - (Your Analyzer)	1	--	--	--	--


**Evaluation Based On - Method wise Peer Grouping**

Your Result	144.00	SDI / z-Score	-0.99
Assigned Value	150.06	Average SDI	-0.78
SDPA	6.100	Uncertainty of AV	0.74
Standard Unit Result Val	144.00	%DEV	-4.16


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Sodium		N	Mean	SD	CV%	Excluded
	All Analyzer	227	149.43	5.03	3.36	4
<b>Your Analyzer :</b>	<b>Allcare AC-9801</b>	1	--	--	--	--
	Fuji Dri-Chem NX500/i	38	149.39	4.55	3.05	2
	Roche cobas c 311	22	149.86	4.53	3.02	--
	Roche AVL 9180	19	149.11	4.73	3.17	--
	Fuji Dri-Chem 4000i	18	151.44	6.30	4.16	--
	Fuji Dri-Chem NX600	14	150.79	2.81	1.86	--
	EasyLyte	10	147.59	3.83	2.60	--
	BeneSphera Electrolyte Analyzer	9	142.29	4.83	3.39	--
	Roche cobas 6000	7	151.57	5.35	3.53	--

Remarks: Interpretation of your reports

- If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Total Protein, g/dL**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	280	6.87	0.42	6.06	9
Biuret - (Your Method)	272	6.87	0.42	6.04	8
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

<b>Your Result</b>	6.78	<b>SDI / z-Score</b>	0.29
<b>Assigned Value</b>	6.68	<b>Average SDI</b>	0.11
<b>SDPA</b>	0.350	<b>Uncertainty of AV</b>	0.03
<b>Standard Unit Result Val</b>	6.78	<b>%DEV</b>	-1.27

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Total Protein		N	Mean	SD	CV%	Excluded
	All Analyzer	280	6.87	0.42	6.06	9
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	38	7.09	0.29	4.03	--
	Roche cobas c 311	29	6.68	0.14	2.04	--
	Erba EM 200	23	7.09	0.41	5.82	--
	Roche cobas c 111	20	6.68	0.39	5.82	--
	Biosystems A15	16	6.77	0.51	7.51	1
	Fuji Dri-Chem 4000i	15	6.93	0.40	5.82	1
	Fuji Dri-Chem NX600	14	6.79	0.22	3.29	1
	Roche cobas INTEGRA	13	6.52	0.17	2.53	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Triglycerides, mg/dl**

**Comparative Statistics**



	N	Mean	SD	CV%	Excluded
All Methods	272	134.52	14.06	10.45	5
Enzymatic, end point - (Your Method)	61	141.77	12.31	8.69	--
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	134.70	SDI / z-Score	-0.15
Assigned Value	136.38	Average SDI	-0.07
SDPA	11.000	Uncertainty of AV	0.78
Standard Unit Result Val	134.70	%DEV	-4.98

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, Triglycerides**

	N	Mean	SD	CV%	Excluded
All Analyzer	272	134.52	14.06	10.45	5
<b>Your Analyzer : Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
Fuji Dri-Chem NX500/i	32	144.49	9.58	6.63	--
Roche cobas c 311	29	136.66	3.76	2.75	--
Erba EM 200	21	116.06	9.39	8.09	--
Roche cobas c 111	21	141.86	11.44	8.06	--
Biosystems A15	16	125.45	20.04	15.98	--
Fuji Dri-Chem 4000i	14	142.93	12.51	8.75	--
Roche cobas INTEGRA	13	136.92	4.48	3.27	--
Fuji Dri-Chem NX600	13	141.00	9.84	6.98	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Urea, mg/dl**



**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	283	20.68	2.11	10.22	14
Urease, UV - (Your Method)	48	20.53	2.50	12.19	5
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--

<b>Your Result</b>	19.70	<b>SDI / z-Score</b>	-0.78
<b>Assigned Value</b>	21.07	<b>Average SDI</b>	-0.76
<b>SDPA</b>	1.761	<b>Uncertainty of AV</b>	0.48
<b>Standard Unit Result Val</b>	19.70	<b>%DEV</b>	-4.04

Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

Biochemistry, Urea		N	Mean	SD	CV%	Excluded
	All Analyzer	283	20.68	2.11	10.22	14
<b>Your Analyzer :</b>	<b>Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
	Fuji Dri-Chem NX500/i	35	21.06	1.20	5.68	--
	Roche cobas c 311	29	20.89	1.19	5.69	1
	Erba EM 200	23	22.09	3.06	13.86	--
	Roche cobas c 111	22	19.63	1.61	8.20	1
	Biosystems A15	16	19.16	3.33	17.39	--
	Fuji Dri-Chem 4000i	15	20.49	0.92	4.49	1
	Fuji Dri-Chem NX600	14	20.76	0.63	3.04	1
	Roche cobas INTEGRA	13	19.55	1.34	6.86	--

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error





**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Biochemistry, Uric Acid, mg/dl**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	275	4.48	0.73	16.24	8
Uricase, colorimetric - (Your Method)	154	4.39	0.66	15.07	3
Beckman Coulter AU680 - (Your Analyzer)	1	--	--	--	--


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	4.19	SDI / z-Score	0.39
Assigned Value	4.06	Average SDI	0.38
SDPA	0.320	Uncertainty of AV	0.02
Standard Unit Result Val	4.19	%DEV	-4.60


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



PC-1022

**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**Absolute Z-Score Average - Across Rounds**



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

**Biochemistry, Uric Acid**

	N	Mean	SD	CV%	Excluded
All Analyzer	275	4.48	0.73	16.24	8
<b>Your Analyzer : Beckman Coulter AU680</b>	<b>1</b>	<b>--</b>	<b>--</b>	<b>--</b>	<b>--</b>
Fuji Dri-Chem NX500/i	33	4.36	0.28	6.51	--
Roche cobas c 311	27	4.06	0.15	3.67	--
Erba EM 200	21	4.96	0.31	6.25	2
Roche cobas c 111	21	4.11	0.26	6.28	1
Fuji Dri-Chem 4000i	18	4.27	0.26	6.02	2
Biosystems A15	16	5.57	0.61	10.95	--
Fuji Dri-Chem NX600	14	4.29	0.15	3.47	--
Roche cobas INTEGRA	13	4.23	0.29	6.86	1

Remarks: Interpretation of your reports

- a) If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- b) If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- c) If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416,  
Commercial Building - 1A,  
Kohinoor Mall, Kirol Road, Kurla (W),  
Mumbai - 400070



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

**HbA1c, HbA1c, %**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	127	4.76	0.33	6.96	6
HPLC - (Your Method)	51	4.73	0.25	5.22	1
Biorad D-10 - (Your Analyzer)	33	4.74	0.19	4.07	2


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	4.70	SDI / z-Score	-0.14
Assigned Value	4.74	Average SDI	0.00
SDPA	0.250	Uncertainty of AV	0.03
Standard Unit Result Val	4.70	%DEV	-0.84


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 2

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

HbA1c, HbA1c		N	Mean	SD	CV%	Excluded
	All Analyzer	127	4.76	0.33	6.96	6
<b>Your Analyzer :</b>	<b>Biorad D-10</b>	<b>33</b>	<b>4.74</b>	<b>0.19</b>	<b>4.07</b>	<b>2</b>
	Tosoh G8 Auto HPLC	16	4.81	0.20	4.20	--
	Tosoh HLC 723 GX	11	4.79	0.15	3.05	--
	Erba EM 200	10	4.43	0.68	15.28	2
	Lifotronic GH-900Plus	10	4.61	0.26	5.55	--
	Mispa i2	9	4.68	0.39	8.33	--

Remarks: Interpretation of your reports

- If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416,  
Commercial Building - 1A,  
Kohinoor Mall, Kirol Road, Kurla (W),  
Mumbai - 400070



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309

Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

**HbA1c, HbA1c, %**

**Comparative Statistics**

	N	Mean	SD	CV%	Excluded
All Methods	128	7.32	0.34	4.68	7
HPLC - (Your Method)	51	7.41	0.28	3.73	3
Biorad D-10 - (Your Analyzer)	33	7.46	0.31	4.16	1


**Evaluation Based On - Equipment wise Peer Grouping**

Your Result	7.50	SDI / z-Score	0.14
Assigned Value	7.45	Average SDI	0.00
SDPA	0.400	Uncertainty of AV	0.03
Standard Unit Result Val	7.50	%DEV	0.55


Note : # - Not evaluated / Bimodal distribution

X : SDI more than 4


Frequency Histogram




Z-Score Trend - Within Round



Frequency Histogram Your Method



Frequency Histogram Your Analyzer



**Analyte Methodwise Result Report**

Lab : 775

Cycle No : 220309


Lab Name : DR AMRISH MEHTA'S APPLE DIAGNOSTIC

Sample No : 1

Report Date : 10/12/2022

Sample Date : 30-09-2022

Absolute Z-Score Average - Across Rounds



Note : SDI with \* marked are not considered for evaluation.

**Comments :**

HbA1c, HbA1c		N	Mean	SD	CV%	Excluded
	All Analyzer	128	7.32	0.34	4.68	7
<b>Your Analyzer :</b>	<b>Biorad D-10</b>	<b>33</b>	<b>7.46</b>	<b>0.31</b>	<b>4.16</b>	<b>1</b>
	Tosoh G8 Auto HPLC	16	7.45	0.18	2.46	--
	Tosoh HLC 723 GX	11	7.31	0.24	3.34	1
	Erba EM 200	10	6.85	0.19	2.70	2
	Lifotronic GH-900Plus	10	7.46	0.34	4.56	--
	Mispa i2	9	7.32	0.30	4.09	--

Remarks: Interpretation of your reports

- If your SDI lies between +/- 2.0 SDI then your results are well within limits.
- If your SDI lies between +/-2.0 to 2.99 SDI then it is a warning alert flag.
- If your SDI is +/- 3.0 SDI or more then it is an action alert flag.

Interpretation when an analyte is reported for multiple samples in a Round:

- Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416,  
Commercial Building - 1A,  
Kohinoor Mall, Kirol Road, Kurla (W),  
Mumbai - 400070

