



CENTRE for HIV TESTING
State Reference Laboratory for HIV
Department of Microbiology
Lady Hardinge Medical College & Associated Hospitals
New Delhi

Ref. No.

Date: 31/10/22

EQAS REPORT (SRL to ICTCs/PPTCTs)

Round: I

Name of Proficiency Testing provider: SRL, CHT-LHMC Date of panel distribution: 21/10/22

Name of the participating laboratory: Acharya Shree Bhikshu Govt. Hospital, ICTC

Date of result received: 26/10/22

Sample code	Standard result	Results obtained by participating laboratory	% matching
A2	Positive for HIV-1 abs	Reactive (HIV-1)	100
A3	Negative for HIV abs	Non-reactive	100
A4	Positive for HIV-1 abs	Reactive (HIV-1)	100
A1	Negative for HIV abs	Non-reactive	100

Remarks: Performance is satisfactory.

Dr. V. S. Randhawa
Name of Laboratory In-Charge:


Signature of Laboratory In-Charge:

*** End of Report***