

Date: 16/11/2022

Effective Date: [REDACTED]

## Certificate of Calibration

**Customer Name:** Redcliffe Lifetech Private Ltd Gwalior

**Model :** Automated Hematology Analyzer Elite 580

**Serial No.:** [REDACTED] K11052203048

**Calibration Done Date:** 16.11.22

**Next Calibration Due Date On or Before:** [REDACTED] 15.11.23

**Lab In-charge:** [REDACTED]

*This is to certify that the above-mentioned product has been verified of calibration for CBC 5 parameters (WBC, RBC, HGB, MCV and PLT) according to the standard procedures provided by Erba Lachema s.r.o, Karasek.*

Calibration at site performed by

Engineer Name Gaurav Chand Gupta

Designation Sr.service engineer

Transasia Bio-Medicals Ltd

Location Gwalior

Encl:

1. Certificate of Inspection
2. Assay Sheet of Hematology Calibrator (H Cal)
3. Printouts
4. Traceability Document