



DELHI CALIBRATION LABORATORY PVT. LTD.

ISO/IEC 17025 : 2017 NABL Accredited Laboratory

House No.-7, Avtar Enclave, Paschim Vihar, New Delhi-110063 (INDIA)

Ph. No.: (011) 35009896, (+91) 9818711781, 9810977012, 9015904300

Email ID: delhicallibrationlaboratory@gmail.com

Website: www.dcllab.in



CC-3444

Calibration Certificate

Page: 01 of 02

Calibration Certificate / ULR No.: CC34442200001610F	Calibration Request No.: DCL/RN-22/0199
Certificate Issued on: 04-11-2022	Date of Receipt: 02-11-2022
Suggested date for the next calibration: 02-11-2023	Date of Calibration: 02-11-2022

CALIBRATED FOR: M/s Origyn Path Lab
4th Floor, HB Twin Tower, Netaji Subhash Place,
Pitampura, New Delhi- 110034

ITEM'S DESCRIPTION:

Name	:	Micro Pipette
Make / Model No.	:	Micropet
Serial No.	:	-----
Identification No.	:	MP- 01
Range	:	20 µl to 200 µl
Least Count	:	1 µl
Location	:	-----
Physical Condition	:	OK

Zero Error: ----

Environmental Condition	Temperature	: 23 °C ± 2 °C
	Relative Humidity	: 40 % to 60 %

STANDARD REFERENCE DETAILS (Traceable to National Standard)

Name of Standard	Make / Model No.	I.D. No.	Certificate No.	Traceable	Calibration Due Date	Uncertainty (At 95% C.L. approx.)
Digital Weighing Balance	Shimadzu / AUW220D	DCL/STD/ DWB-01	PV/2204/ 0276	PV Calibration (Ghaziabad)	20-04-2023	± 0.12 mg (k=2)

Equipments Used: Thermometer, Thermo-Hygrometer, Barometer, Distilled Water & Glasswares

Principle / Methodology of Calibration: As per Calibration Procedure No.: DCL/CP/M-65

Used Standard: IS/ISO: 8655 & ISO/TR 20461:2000 (E)

CALIBRATION RESULTS (IN-LAB)

Average U.U.C. Value (µl)	Average Measured Reference Standard value (µl)	Expanded Uncertainty [^] (µl)
20	19.94	± 0.120
100	99.92	± 0.120
200	199.85	± 0.357

Calibrated by: Dayaram Sushant
(Sr. Calibration Engineer)



Approved by: B.G. Sharma
(Technical Head)

- Note :**
- This certificate refers only to the particular item submitted for calibration.
 - This certificate shall not be reproduced except in full without the written approval of Laboratory.
 - Reported Results are valid at the time of and under the stated conditions of measurement.
 - Laboratory Standards are traceable to National Standards.
 - In case any reconfirmation of contents of this calibration certificate is required, please contact our office.



DELHI CALIBRATION LABORATORY PVT. LTD.

Format No. : QPM-32F



ISO/IEC 17025 : 2017 NABL Accredited Laboratory

House No.-7, Avtar Enclave, Paschim Vihar, New Delhi-110063 (INDIA)

Ph. No.: (011) 35009896, (+91) 9818711781, 9810977012, 9015904300

Email ID: delhicallbrationlaboratory@gmail.com

Website: www.dcllab.in

CC-3444

Calibration Certificate

Page: 02 of 02

Calibration Certificate / ULR No.: CC34442200001610F	Calibration Request No.: DCL/RN-22/0199
Certificate Issued on: 04-11-2022	Date of Receipt: 02-11-2022
Suggested date for the next calibration: 02-11-2023	Date of Calibration: 02-11-2022


Remarks:

- 1- Expanded Uncertainty[^] of measurement evaluated at a confidence Level of approximately 95% with Coverage factor k=2 for a normal probability distribution.
- 1- U.U.C. defined above as Unit Under Calibration.
- 2- The reported volume is at 27 °C.
- 3- Suggested date for the next calibration mentioned as per customer requisition.
- 4- Decision Rule: Not Required




End of Report



Calibrated by: 
Dayaram Sushant
(Sr. Calibration Engineer)



Approved by: 
B.G. Sharma
(Technical Head)

Note :

1. This certificate refers only to the particular Item submitted for calibration.
2. This certificate shall not be reproduced except in full without the written approval of Laboratory.
3. Reported Results are valid at the time of and under the stated conditions of measurement.
4. Laboratory Standards are traceable to National Standards.
5. In case any reconfirmation of contents of this calibration certificate is required, please contact our office.