

## VIEW LAB MONTHLY SUMMARY

Lab Name FAMILY HEALTH CENTRE  
 Lab No 9960  
 Month December  
 Year 2022  
 Constituent Group Chemistry I

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 14/12/2022

Date of Report Published : 05/01/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Agape	183	180.78	5.46	9.87	220.72 mg/dL	4.05	1.46
2	UREA	Urease UV / GLDH	Agape	188	99.01	6.76	6.69	107.96 mg/dL	1.34	0.98
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Agape	104	2.90	9.24	0.27	3.04 mg/dL	0.52	0.05
4	T.BILIRUBIN	Diazonium salt ( Colorimetric ) / Jendrassik	Agape	167	3.10	10.02	0.31	3.75 mg/dL	2.09	0.05
5	T-PROTEIN	Biuret - Colorimetric	Agape	199	5.14	6.33	0.32	5.04 g/dL	-0.31	0.05
6	ALBUMIN	BCG - colorimetric	Agape	190	3.37	5.22	0.18	3.08 g/dL	-1.65	0.03
7	URIC ACID	Enzymatic / Uricase Colorimetric	Agape	187	6.93	8.40	0.58	7.85 mg/dL	1.58	0.09
8	CHOLESTEROL	CHOD-PAP	Agape	210	99.42	6.69	6.65	110.86 mg/dL	1.72	0.92
9	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Agape	199	173.78	6.65	11.56	165.09 mg/dL	-0.75	1.64
10	HDL	Direct method / Enzymatic colorimetric	Agape	169	24.88	10.69	2.66	25.96 mg/dL	0.41	0.41
11	AST	UV kinetic(with & without PLP (P-5-P))	Agape	195	59.86	8.21	4.91	59.01 U/L	-0.17	0.70
12	ALT	UV kinetic(with & without PLP (P-5-P))	Agape	167	96.58	10.78	10.41	68.2 U/L	-2.73	1.61
13	ALP	PNP AMP kinetic	Agape	71	78.77	16.76	13.20	123.89 U/L	3.42	3.13

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

[Modify Address](#)

[Modify Methods](#)

[Submit Lab Results](#)

[View Lab Monthly Summary](#)