



LAURA: Calibration procedure

To,
Redcliffe Lifetech Pvt. Limited.
F82, Ground Floor, Patel Nagar,
City Centre, Opp. Osho Meditation Centre,
Gwalior

Instrument Name : Laura
Instrument Sr. No. : 0844-1-202110

We Erba Lachema s.r.o. with registered office in Czech Republic, Karásek 2219/1D, 621 00 Brno, represented by its Product Manager Veronika Plešková hereby confirm as producer of LAURA instrument that this device is firmly calibrated by manufacturer and does not require any regular calibration by the user. Correct calibration of the instrument is controlled by grey strips provided initially together with the instrument.


Veronika Plešková



Product manager, Urine

Analysis Erba Corporate

Service Ltd.
GW1 Great West House
Great West Road, London TW8 9DF



Erba Corporate Services Ltd . Registered in England and Wales . Company No:
11100097 Registered Office: c/o Erba Molecular, Unit 4 Bartholemews, Ely,
Cambridgeshire, CB74EA, UK



2023.01.11

09:05

DEKAPHAN LAURA

Seq.No: 0001

URINORM N

2023.01.11

09:34

.....
BLD NEG
BIL NEG
UBG NORM
KET NEG
GLU NEG
PRO NEG
pH 6
NIT NEG
LEU NEG
SG 1.025

QC TEST: OK
<Admin>

Shirwan
11/1/23

DEKAPHAN LAURA

Seq.No: 0002

URINORM P

2023.01.11

09:36

.....
*BLD 3+
*BIL 3+
*UBG 2+
*KET 3+
*GLU 2+
*PRO 3+
pH 9
*NIT POS
*LEU 500 Leu/u1
SG 1.000

QC TEST: OK
<Admin>

