




LAURA: Calibration procedure

To,
Redcliffe Lifetech Pvt. Limited.
Tilak Nagar, Delhi.

Instrument Name : Laura V2

Instrument Sr. No. : 0959-1-202201

We Erba Lachema s.r.o. with registered office in Czech Republic, Karásek 2219/1D, 621 00 Brno, represented by its Product Manager Veronika Plešková hereby confirm as producer of LAURA instrument that this device is firmly calibrated by manufacturer and does not require any regular calibration by the user. Correct calibration of the instrument is controlled by grey strips provided initially together with the instrument.


Veronika Plešková



Product manager, Urine

Analysis Erba Corporate

Service Ltd.

GW1 Great West House

Great West Road, London TW8 9DF



Erba Corporate Services Ltd . Registered in England and Wales . Company No:
11100097 Registered Office: c/o Erba Molecular, Unit 4 Bartholemews, Ely,
Cambridgeshire, CB74EA, UK



Urinalysis

LAURA

12.01.2023

09:33 AM

DEKAPHAN LAURA

Seq.No: 0001

URINORM N

12.01.2023

09:38 AM

.....

BLD	NEG
BIL	NEG
UBG	NORM
KET	NEG
GLU	NEG
PRO	NEG
pH	6
NIT	NEG
LEU	NEG
SG	1.025

QC TEST: OK

<Admin>

Keeler
12/1/2023

DEKAPHAN LAURA

Seq.No: 0002

URINORM P

12.01.2023

09:40 AM

.....

*BLD	50	Ery/u1
*BIL	3	mg/dl
*UBG	3	mg/dl
*KET	52	mg/dl
*GLU	500	mg/dl
*PRO	500	mg/dl
pH	9	
*NIT	POS	
*LEU	125	Leu/u1
SG	1.000	

QC TEST: OK

<Admin>

Admin
12/1/23